



TOTAL VOLATILE SOLIDS 160.4

Run Number: LB117415

SUPERVISOR: apatel

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

TEMP1 IN: 104 °C 11/17/2021 16:00 TEMP1 OUT: 104 °C 11/18/2021 08:00
TEMP2 IN: 103 °C 11/18/2021 08:30 TEMP2 OUT: 103 °C 11/18/2021 10:00
TEMP3 IN: 540 °C 11/18/2021 10:30 TEMP3 OUT: 540 °C 11/18/2021 12:00
TEMP4 IN: 560 °C 11/18/2021 12:30 TEMP4 OUT: 560 °C 11/18/2021 14:00

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Empty Dish + Sample Weight (g)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (%)
1	LB117415BL	74.8397	74.8397	74.8397	74.8397	74.8397	74.8397	74.8397	0.0000	0
2	M4731-01	55.8883	55.8883	75.9421	70.1171	70.1171	69.3139	69.3139	0.8032	5.6
3	M4731-01DUP	58.3188	58.3188	80.1371	73.8593	73.8593	72.9978	72.9978	0.8615	5.5

A = Sample Weight (g)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{F - E} \star 100$

11/17/21

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS M4731 S WorkList ID : 154435 Department : Wet-Chemistry Date : 11-17-2021 13:11:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/22/2021	Solid	M4731-01	TVS	Cool 4 deg C	TULL02	D32	C-1600-FLOOD-WALL-JC	11/17/2021	160.4

Date/Time 11-17-21 15:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JD CSM

Date/Time 11-17-21 16:30
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JP (WC)