

LB117417


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44267



A= 0.0000e+000

B= 2.2824e-004 *slope*C= -1.2731e-001 *y-intercept*

Rho= 0.9999330

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.011	0.011	604	0.000	604					
0.05	0.050					205	0				
0.20	0.200	0.219	0.019	1517	0.0 %	1517					
2.50	2.500	2.476	-0.024	11407	0.0 %	11407					
5.00	5.000	4.941	-0.059	22207	0.0 %	22207					
7.50	7.500	7.577	0.077	33754	0.0 %	33754					
10.0	10.000	9.976	-0.024	44267	0.0 %	44267					

10
 -1
 -1
 1
 0

LB117417
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	604	-		0 SFAM01.1	PPB	11/17/2021 15:00	Std
	0.2 S01	1517	-		0.2 SFAM01.1	PPB	11/17/2021 15:05	Std
	2.5 S02	11407	-		2.5 SFAM01.1	PPB	11/17/2021 15:07	Std
	5 S03	22207	-		5 SFAM01.1	PPB	11/17/2021 15:09	Std
	7.5 S04	33754	-		7.5 SFAM01.1	PPB	11/17/2021 15:11	Std
	10 S05	44267	-		10 SFAM01.1	PPB	11/17/2021 15:14	Std
ICV088	ICV088	16897	3.7293 -		SFAM01.1	PPB	11/17/2021 15:17	SMPL
ICB088	ICB088	573	0.0035 -		SFAM01.1	PPB	11/17/2021 15:19	SMPL
CCV078	CCV078	22240	4.9488 -		SFAM01.1	PPB	11/17/2021 15:21	SMPL
CCB078	CCB078	572	0.0032 -		SFAM01.1	PPB	11/17/2021 15:23	SMPL
PB140842BL	PBS842	608	0.0115 -		SFAM01.1	PPB	11/17/2021 15:26	SMPL
M4486-01	GB7X3	817	0.0592 -		SFAM01.1	PPB	11/17/2021 15:28	SMPL
M4486-02	GB7X4	660	0.0233 -		SFAM01.1	PPB	11/17/2021 15:30	SMPL
M4486-03	GB7X5	1173	0.1404 -		SFAM01.1	PPB	11/17/2021 15:32	SMPL
M4486-04	GB7X6	1070	0.1169 -		SFAM01.1	PPB	11/17/2021 15:35	SMPL
M4486-05	GB7X7	646	0.0201 -		SFAM01.1	PPB	11/17/2021 15:37	SMPL
M4486-06	GB7X8	8230	1.7511 -		SFAM01.1	PPB	11/17/2021 15:39	SMPL
M4486-07	GB7X9	688	0.0297 -		SFAM01.1	PPB	11/17/2021 15:41	SMPL
M4486-08	GB7Y0	657	0.0226 -		SFAM01.1	PPB	11/17/2021 15:44	SMPL
M4486-09	GB7Y1	676	0.027 -		SFAM01.1	PPB	11/17/2021 15:46	SMPL
M4486-10	GB7Y2	634	0.0174 -		SFAM01.1	PPB	11/17/2021 15:48	SMPL
M4486-11	GB7Y3	2860	0.5255 -		SFAM01.1	PPB	11/17/2021 15:50	SMPL
M4486-12	GB7Y4	2947	0.5453 -		SFAM01.1	PPB	11/17/2021 15:53	SMPL
M4486-13	GB7Y5	2399	0.4202 -		SFAM01.1	PPB	11/17/2021 15:55	SMPL
M4486-14	GB7Y6	1944	0.3164 -		SFAM01.1	PPB	11/17/2021 15:57	SMPL
M4486-15	GB7Y7	2682	0.4848 -		SFAM01.1	PPB	11/17/2021 15:59	SMPL
M4486-16	GB7Y7D	2968	0.5501 -		SFAM01.1	PPB	11/17/2021 16:02	SMPL
M4486-17	GB7Y7S	13954	3.0576 -		SFAM01.1	PPB	11/17/2021 16:04	SMPL
M4486-18	GB7Y8	21370	4.7502 -		SFAM01.1	PPB	11/17/2021 16:09	SMPL
CCV079	CCV079	21288	4.7315 -		SFAM01.1	PPB	11/17/2021 16:13	SMPL
CCB079	CCB079	517	-0.0093 -		SFAM01.1	PPB	11/17/2021 16:15	SMPL