



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117423

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	11/17/2021	17:15
2	CAL2	1	Water	NA	NA	20.9	7.00	11/17/2021	17:16
3	CAL3	1	Water	NA	NA	20.9	10.03	11/17/2021	17:17
4	ICV	1	Water	NA	NA	20.8	7.00	11/17/2021	17:19
5	CCV1	1	Water	NA	NA	20.8	2.01	11/17/2021	17:20
6	M4729-01	1	Solid	20.03	20	20.8	5.56	11/17/2021	17:22
7	M4731-02	1	Solid	20.04	20	20.3	11.79	11/17/2021	17:25
8	M4731-02DUP	1	Solid	20.03	20	20.4	11.80	11/17/2021	17:26
9	M4736-02	1	Solid	20.04	20	20.6	6.22	11/17/2021	17:30
10	CCV2	1	Water	NA	NA	20.8	12.03	11/17/2021	17:31

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M4731 WorkList ID : 154440 Department : Wet-Chemistry Date : 11-17-2021 13:14:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/17/2021	Solid	M4729-01	Corrosivity	Cool 4 deg C	PSEG03	D32	EIGI-284	11/17/2021	9045D
11/17/2021	Solid	M4731-02	Corrosivity	Cool 4 deg C	TULL02	D32	C-1600-FLOOD-WALL-JC	11/17/2021	9045D
11/17/2021	Solid	M4736-02	Corrosivity	Cool 4 deg C	PSEG03	D32	VNJ-202	11/17/2021	9045D

Date/Time 11-17-21 17:00
 Raw Sample Received by: JP WCI
 Raw Sample Relinquished by: CP SM

Date/Time 11-17-21 18:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP WCI