



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117424

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	11/17/2021	17:15
2	CAL2	1	Water	NA	NA	20.9	7.00	11/17/2021	17:16
3	CAL3	1	Water	NA	NA	20.9	10.03	11/17/2021	17:17
4	ICV	1	Water	NA	NA	20.8	7.00	11/17/2021	17:19
5	CCV1	1	Water	NA	NA	20.8	2.01	11/17/2021	17:20
6	M4731-01	1	Solid	20.03	20	20.8	11.79	11/17/2021	17:25
7	M4731-01DUP	1	Solid	20.03	20	20.4	11.80	11/17/2021	17:26
8	M4732-01	1	Solid	20.03	20	21.5	5.17	11/17/2021	17:33
9	M4732-03	1	Solid	20.04	20	20.8	4.92	11/17/2021	17:35
10	M4732-05	1	Solid	20.01	20	20.9	4.83	11/17/2021	17:40
11	M4732-07	1	Solid	20.04	20	20.8	6.49	11/17/2021	17:42
12	CCV2	1	Water	NA	NA	20.8	12.03	11/17/2021	17:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4731 S WorkList ID : 154439 Department : Wet-Chemistry Date : 11-17-2021 13:14:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/17/2021	Solid	M4731-01	pH	Cool 4 deg C	TULL02	D32	C-1600-FLOOD-WALL-JC	11/17/2021	9045D
11/17/2021	Solid	M4732-01	pH	Cool 4 deg C	PSEG03	D32	33536	11/17/2021	9045D
11/17/2021	Solid	M4732-03	pH	Cool 4 deg C	PSEG03	D32	33537	11/17/2021	9045D
11/17/2021	Solid	M4732-05	pH	Cool 4 deg C	PSEG03	D32	33538	11/17/2021	9045D
11/17/2021	Solid	M4732-07	pH	Cool 4 deg C	PSEG03	D32	33175	11/17/2021	9045D

Date/Time 11-17-21 17:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CL SM

Date/Time 11-17-21 18:30
 Raw Sample Received by: CL SM
 Raw Sample Relinquished by: JP (WC)