

LB117431


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35445



A= 0.0000e+000

B= 2.8053e-004

C= -2.1308e-002

Rho= 0.9998795

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.008	-0.008	46	0.000	46					1
0.2	0.200	0.202	0.002	795	0.0 %	795					-2
2.5	2.500	2.442	-0.058	8781	0.0 %	8781					1
5.0	5.000	5.066	0.066	18135	0.0 %	18135					1
7.5	7.500	7.576	0.076	27083	0.0 %	27083					1
10.0	10.000	9.922	-0.078	35445	0.0 %	35445					1

LB117431 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Nov 2021 11:28:27

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	SD	46	-	0.00007470A	PPB	18 Nov 2021 11:32:19	S	Std
0.2 - 1	SD 2	795	-	0.20007470A	PPB	18 Nov 2021 11:34:36	S	Std
2.5 - 1	SD 5	8781	-	2.50007470A	PPB	18 Nov 2021 11:36:53	S	Std
5.0 - 1	SD 5	18135	-	5.00007470A	PPB	18 Nov 2021 11:39:10	S	Std
7.5 - 1	SD 5	27083	-	7.50007470A	PPB	18 Nov 2021 11:41:26	S	Std
10.0 - 1	SD 10	35445	-	10.00007470A	PPB	18 Nov 2021 11:46:15	S	Std
ICV86 - 1	ICV86	13125	3.6607	-7470A	PPB	18 Nov 2021 11:48:54	U	SMPL
ICB86 - 1	ICB86	-18	-0.0264	-7470A	PPB	18 Nov 2021 11:51:09	U	SMPL
CCV80 - 1	CCV80	18241	5.0959	-7470A	PPB	18 Nov 2021 11:53:24	U	SMPL
CCB80 - 1	CCB80	-34	-0.0308	-7470A	PPB	18 Nov 2021 11:55:39	U	SMPL
CRA - 1	CRA	781	0.1978	-7470A	PPB	18 Nov 2021 11:57:54	U	SMPL
HighStd - 1	HighStd	35742	10.0055	-7470A	PPB	18 Nov 2021 12:00:09	U	SMPL
ChkStd - 1	ChkStd	24048	6.7249	-7470A	PPB	18 Nov 2021 12:02:24	U	SMPL
PB140853BL - 1	PBW	-95	-0.0480	-7470A	PPB	18 Nov 2021 12:04:41	U	SMPL
PB140853BS - 1	LCBW	12590	3.5106	-7470A	PPB	18 Nov 2021 12:06:56	U	SMPL
M4700-02 - 1	IMPOUND-LOT-SP-JC-11-15-21	-36	-0.0314	-7470A	PPB	18 Nov 2021 12:09:12	U	SMPL
M4707-10 - 1	1-COMP	4	-0.0202	-7470A	PPB	18 Nov 2021 12:11:29	U	SMPL
M4707-19 - 1	2-COMP	14	-0.0174	-7470A	PPB	18 Nov 2021 12:13:46	U	SMPL
M4707-30 - 1	3-COMP	10	-0.0185	-7470A	PPB	18 Nov 2021 12:16:04	U	SMPL
M4707-40 - 1	4-COMP	15	-0.0171	-7470A	PPB	18 Nov 2021 12:18:19	U	SMPL
CCV81 - 1	CCV81	18025	5.0353	-7470A	PPB	18 Nov 2021 12:20:34	U	SMPL
CCB81 - 1	CCB81	-30	-0.0297	-7470A	PPB	18 Nov 2021 12:22:49	U	SMPL
M4713-03 - 1	TP-2-COMP	-1	-0.0216	-7470A	PPB	18 Nov 2021 12:25:04	U	SMPL
M4713-03DUP - 1	TP-2-COMPDUP	4	-0.0202	-7470A	PPB	18 Nov 2021 12:27:19	U	SMPL
M4713-03MS - 1	TP-2-COMPMS	12871	3.5894	-7470A	PPB	18 Nov 2021 12:29:35	U	SMPL
M4713-03MSD - 1	TP-2-COMPMSD	12615	3.5176	-7470A	PPB	18 Nov 2021 12:31:51	U	SMPL
PB140853TB - 1	PB1408816TB	-51	-0.0356	-7470A	PPB	18 Nov 2021 12:34:07	U	SMPL
M4713-03LX5 - 1		21	-0.0154	-7470A	PPB	18 Nov 2021 12:36:23	U	SMPL
M4713-03A - 1		14380	4.0127	-7470A	PPB	18 Nov 2021 12:38:40	U	SMPL
CCV82 - 1	CCV82	17837	4.9825	-7470A	PPB	18 Nov 2021 12:40:57	U	SMPL
CCB82 - 1	CCB82	-15	-0.0255	-7470A	PPB	18 Nov 2021 12:43:16	U	SMPL