



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117432

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/18/2021	09:20
2	CAL2	1	Water	NA	NA	20.3	7.00	11/18/2021	09:21
3	CAL3	1	Water	NA	NA	20.3	10.03	11/18/2021	09:22
4	ICV	1	Water	NA	NA	20.4	7.00	11/18/2021	09:25
5	CCV1	1	Water	NA	NA	20.4	2.01	11/18/2021	09:30
6	M4730-01	1	Water	NA	NA	20.4	7.13	11/18/2021	09:32
7	M4730-02	1	Water	NA	NA	20.5	7.10	11/18/2021	09:35
8	M4732-11	1	Water	NA	NA	20.3	5.95	11/18/2021	09:40
9	M4733-03	1	Water	NA	NA	20.8	7.62	11/18/2021	09:45
10	M4733-04	1	Water	NA	NA	20.4	7.54	11/18/2021	09:49
11	M4733-05	1	Water	NA	NA	20.5	7.30	11/18/2021	09:51
12	M4737-08	1	Water	NA	NA	21.5	6.98	11/18/2021	10:00
13	M4737-09	1	Water	NA	NA	21.6	6.88	11/18/2021	10:02
14	M4737-10	1	Water	NA	NA	21.4	6.92	11/18/2021	10:05
15	M4737-11	1	Water	NA	NA	21.5	6.88	11/18/2021	10:06
16	CCV2	1	Water	NA	NA	20.4	12.02	11/18/2021	10:10
17	M4737-12	1	Water	NA	NA	21.3	6.94	11/18/2021	10:15
18	M4737-13	1	Water	NA	NA	21.4	6.94	11/18/2021	10:19
19	M4737-14	1	Water	NA	NA	21.2	7.06	11/18/2021	10:22
20	M4737-15	1	Water	NA	NA	21.4	6.91	11/18/2021	10:25

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M4737-16	1	Water	NA	NA	21.3	6.85	11/18/2021	10:26
22	M4737-17	1	Water	NA	NA	21.3	6.94	11/18/2021	10:30
23	M4737-18	1	Water	NA	NA	21.4	6.99	11/18/2021	10:33
24	M4737-19	1	Water	NA	NA	21.5	6.92	11/18/2021	10:35
25	M4737-20	1	Water	NA	NA	21.7	6.69	11/18/2021	10:41
26	M4737-21	1	Water	NA	NA	21.5	6.54	11/18/2021	10:45
27	CCV3	1	Water	NA	NA	20.9	2.01	11/18/2021	10:48
28	M4737-22	1	Water	NA	NA	21.6	6.66	11/18/2021	10:50
29	M4737-22DUP	1	Water	NA	NA	21.7	6.67	11/18/2021	10:51
30	CCV4	1	Water	NA	NA	20.9	12.03	11/18/2021	11:00

VB117432

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S M4517

WorkList ID : 154445

Department : Wet-Chemistry

Date : 11-18-2021 09:10:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/17/2021	Water	M4730-01	pH	Cool 4 deg C	PSEG03	D32	33520	11/17/2021	9040C
11/17/2021	Water	M4730-02	pH	Cool 4 deg C	PSEG03	D32	33521	11/17/2021	9040C
11/17/2021	Water	M4732-11	pH	Cool 4 deg C	PSEG03	D32	33535	11/17/2021	9040C
11/17/2021	Water	M4733-03	pH	Cool 4 deg C	PSEG03	D32	33518	11/17/2021	9040C
11/17/2021	Water	M4733-04	pH	Cool 4 deg C	PSEG03	D32	33552	11/17/2021	9040C
11/17/2021	Water	M4733-05	pH	Cool 4 deg C	PSEG03	D32	33553	11/17/2021	9040C
11/17/2021	Water	M4737-20	pH	Cool 4 deg C	PSEG03	D32	20924	11/17/2021	9040C
11/17/2021	Water	M4737-21	pH	Cool 4 deg C	PSEG03	D32	20922	11/17/2021	9040C
11/17/2021	Water	M4737-22	pH	Cool 4 deg C	PSEG03	D32	20923	11/17/2021	9040C
11/17/2021	Water	M4737-14	pH	Cool 4 deg C	PSEG03	D32	20916	11/17/2021	9040C
11/17/2021	Water	M4737-15	pH	Cool 4 deg C	PSEG03	D32	20917	11/17/2021	9040C
11/17/2021	Water	M4737-16	pH	Cool 4 deg C	PSEG03	D32	20918	11/17/2021	9040C
11/17/2021	Water	M4737-17	pH	Cool 4 deg C	PSEG03	D32	20919	11/17/2021	9040C
11/17/2021	Water	M4737-18	pH	Cool 4 deg C	PSEG03	D32	20920	11/17/2021	9040C
11/17/2021	Water	M4737-19	pH	Cool 4 deg C	PSEG03	D32	20921	11/17/2021	9040C
11/17/2021	Water	M4737-08	pH	Cool 4 deg C	PSEG03	D32	20915	11/17/2021	9040C
11/17/2021	Water	M4737-09	pH	Cool 4 deg C	PSEG03	D32	20910	11/17/2021	9040C
11/17/2021	Water	M4737-10	pH	Cool 4 deg C	PSEG03	D32	20911	11/17/2021	9040C
11/17/2021	Water	M4737-11	pH	Cool 4 deg C	PSEG03	D32	20912	11/17/2021	9040C
11/17/2021	Water	M4737-12	pH	Cool 4 deg C	PSEG03	D32	20913	11/17/2021	9040C
11/17/2021	Water	M4737-13	pH	Cool 4 deg C	PSEG03	D32	20914	11/17/2021	9040C

Date/Time 11-18-21 09:15
 Raw Sample Received by: JP (WDC)
 Raw Sample Relinquished by: JP (WDC)

Date/Time 11-18-21 12:00
 Raw Sample Received by: JP (WDC)
 Raw Sample Relinquished by: JP (WDC)