



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117440

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	11/18/2021	12:58
2	M4730-03		1	10	100.00	>212.0	>212.0	11/18/2021	13:20
3	M4730-03DUP		1	10	100.00	>212.0	>212.0	11/18/2021	13:45
4	M4732-11		1	15	100.00	>212.0	>212.0	11/18/2021	14:10
5	M4733-03		1	14	100.00	>212.0	>212.0	11/18/2021	14:35
6	M4733-04		1	12	100.00	>212.0	>212.0	11/18/2021	15:05
7	M4733-05		1	13	100.00	>212.0	>212.0	11/18/2021	15:32
8	M4737-04		1	15	100.00	>212.0	>212.0	11/18/2021	15:58
9	M4737-05		1	13	100.00	>212.0	>212.0	11/18/2021	16:20
10	M4737-06		1	14	100.00	>212.0	>212.0	11/18/2021	16:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB117440

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp111721 WorkList ID : 154418 Department : Wet-Chemistry Date : 11-17-2021 09:05:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/24/2021	Water	M4730-03	Flash Point	Cool 4 deg C	PSEG03	D32	33520-21	11/17/2021	1010B
11/24/2021	Water	M4732-11	Flash Point	Cool 4 deg C	PSEG03	D32	33535	11/17/2021	1010B
11/24/2021	Water	M4733-03	Flash Point	Cool 4 deg C	PSEG03	D32	33518	11/17/2021	1010B
11/24/2021	Water	M4733-04	Flash Point	Cool 4 deg C	PSEG03	D32	33552	11/17/2021	1010B
11/24/2021	Water	M4733-05	Flash Point	Cool 4 deg C	PSEG03	D32	33553	11/17/2021	1010B
11/24/2021	Water	M4737-04	Flash Point	Cool 4 deg C	PSEG03	D32	COMP-1A-20910-15	11/17/2021	1010B
11/24/2021	Water	M4737-05	Flash Point	Cool 4 deg C	PSEG03	D32	COMP-2A-20916-21	11/17/2021	1010B
11/24/2021	Water	M4737-06	Flash Point	Cool 4 deg C	PSEG03	D32	COMP-3A-20922-24	11/17/2021	1010B

Date/Time 11/18/21 09:00
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: MB (WC)

Date/Time 11/18/21 16:55
Raw Sample Received by: MB (WC)
Raw Sample Relinquished by: MB (WC)