



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117444

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/18/2021	15:15
2	CAL2	1	Water	NA	NA	20.5	7.01	11/18/2021	15:16
3	CAL3	1	Water	NA	NA	20.2	10.02	11/18/2021	15:17
4	ICV	1	Water	NA	NA	20.3	7.00	11/18/2021	15:18
5	CCV1	1	Water	NA	NA	20.3	2.01	11/18/2021	15:20
6	M4756-04	1	Water	NA	NA	21.6	7.46	11/18/2021	15:22
7	M4756-04DUP	1	Water	NA	NA	21.5	7.48	11/18/2021	15:23
8	M4756-05	1	Water	NA	NA	20.7	7.51	11/18/2021	15:24
9	M4757-01	1	Water	NA	NA	20.5	5.34	11/18/2021	15:25
10	CCV2	1	Water	NA	NA	20.5	12.02	11/18/2021	15:26

WORKLIST(Hardcopy Internal Chain)

VB 11-18-21

WorkList Name : PH W M4737 WorkList ID : 154479 Department : Wet-Chemistry Date : 11-18-2021 12:52:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/18/2021	Water	M4756-04	pH	Cool 4 deg C	PSEG03	D11	DRUM-1	11/18/2021	9040C
11/18/2021	Water	M4756-05	pH	Cool 4 deg C	PSEG03	D11	DRUM-2	11/18/2021	9040C
11/10/2021	Water	M4757-01	pH	Cool 4 deg C	STER17	D11	JS9600	11/10/2021	9040C

Date/Time 11-18-21 15:10
 Raw Sample Received by: JP (WOC)
 Raw Sample Relinquished by: CP SM

Date/Time 11-18-21 17:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WOC)