



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117446

Slope : 98.4

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/18/2021	15:15
2	CAL2	1	Water	NA	NA	20.5	7.01	11/18/2021	15:16
3	CAL3	1	Water	NA	NA	20.2	10.02	11/18/2021	15:17
4	ICV	1	Water	NA	NA	20.3	7.00	11/18/2021	15:18
5	CCV1	1	Water	NA	NA	21.4	2.01	11/18/2021	15:35
6	M4765-03	1	Solid	20	20	21.5	9.80	11/18/2021	15:37
7	M4765-03DUP	1	Solid	20	20	21.4	9.82	11/18/2021	15:38
8	CCV2	1	Water	NA	NA	21.2	12.1	11/18/2021	15:40

WB 117446

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4765 WorkList ID : 154505 Department : Wet-Chemistry Date : 11-18-2021 15:04:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/18/2021	Solid	M4765-03	Corrosivity	Cool 4 deg C	TSLA01	D32	WC-2	11/18/2021	9045D

Date/Time 11-18-21 15:10
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: CP SM

Date/Time 11-18-21 17:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: CP SM