


7470A

LB117457
INSTRUMENT ID: CY1

Linear

 μ Abs.:

42863



A= 0.0000e+000

B= 2.3464e-004

C= 8.8374e-003

Rho= 0.9996484

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	24	0.000	24					9
0.2	0.200	0.218	0.018	893	0.0 %	893					-3
2.5	2.500	2.426	-0.074	10301	0.0 %	10301					3
5.0	5.000	5.140	0.140	21870	0.0 %	21870					-2
7.5	7.500	7.335	-0.165	31222	0.0 %	31222					1
10.0	10.000	10.066	0.066	42863	0.0 %	42863					

LB117457 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Nov 2021 10:55:30					
Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7470A	PPB	19 Nov 2021 10:59:12	S	Std
0.2 - 1	50.2	893	-	0.2000	7470A	PPB	19 Nov 2021 11:01:29	S	Std
2.5 - 1	52.5	10301	-	2.5000	7470A	PPB	19 Nov 2021 11:03:46	S	Std
5.0 - 1	55	21870	-	5.0000	7470A	PPB	19 Nov 2021 11:06:03	S	Std
7.5 - 1	57.5	31222	-	7.5000	7470A	PPB	19 Nov 2021 11:08:20	S	Std
10.0 - 1	510	42863	-	10.0000	7470A	PPB	19 Nov 2021 11:13:12	S	Std
ICV87 - 1	ICV87	17667	4.1542	-	7470A	PPB	19 Nov 2021 11:18:25	U	SMPL
ICB87 - 1	ICB87	-72	-0.0081	-	7470A	PPB	19 Nov 2021 11:20:40	U	SMPL
CCV86 - 1	CCV86	21232	4.9907	-	7470A	PPB	19 Nov 2021 11:22:55	U	SMPL
CCB86 - 1	CCB86	-55	-0.0041	-	7470A	PPB	19 Nov 2021 11:25:10	U	SMPL
CRA - 1	CRA	877	0.2146	-	7470A	PPB	19 Nov 2021 11:27:24	U	SMPL
HighStd - 1	HighStd	41700	9.7933	-	7470A	PPB	19 Nov 2021 11:29:39	U	SMPL
ChkStd - 1	ChkStd	26665	6.2655	-	7470A	PPB	19 Nov 2021 11:31:55	U	SMPL
PB140887BL - 1	PBW	-88	-0.0118	-	7470A	PPB	19 Nov 2021 11:34:11	U	SMPL
PB140887BS - 1	LCSW	13705	3.2246	-	7470A	PPB	19 Nov 2021 11:36:28	U	SMPL
M4729-01 - 1	EIGI-284	-31	0.0016	-	7470A	PPB	19 Nov 2021 11:38:44	U	SMPL
M4731-02 - 1	C-1600-FLOOD-WALL-JG	3	0.0095	-	7470A	PPB	19 Nov 2021 11:41:00	U	SMPL
M4736-02 - 1	VNU-202	25	0.0147	-	7470A	PPB	19 Nov 2021 11:43:18	U	SMPL
M4738-02 - 1	P001-B02-PC001-01	51	0.0208	-	7470A	PPB	19 Nov 2021 11:45:35	U	SMPL
M4738-04 - 1	P001-B02-PC002-01	20	0.0135	-	7470A	PPB	19 Nov 2021 11:47:50	U	SMPL
CCV87 - 1	CCV87	20986	4.9330	-	7470A	PPB	19 Nov 2021 11:50:06	U	SMPL
CCB87 - 1	CCB87	-50	-0.0029	-	7470A	PPB	19 Nov 2021 11:52:21	U	SMPL
M4738-06 - 1	P001-B02-PC003-01	30	0.0159	-	7470A	PPB	19 Nov 2021 11:54:37	U	SMPL
M4738-08 - 1	P001-B03-PC001-01	26	0.0149	-	7470A	PPB	19 Nov 2021 11:56:52	U	SMPL
M4738-08DUP - 1	P001-B03-PC001-01DUP	27	0.0152	-	7470A	PPB	19 Nov 2021 11:59:07	U	SMPL
M4738-11 - 1	M4738-08MS	14826	3.4876	-	7470A	PPB	19 Nov 2021 12:03:54	U	SMPL
M4738-12 - 1	M4738-08MS	15133	3.5596	-	7470A	PPB	19 Nov 2021 12:06:09	U	SMPL
M4738-14 - 1	P001-B03-PC001-02	-27	0.0025	-	7470A	PPB	19 Nov 2021 12:08:25	U	SMPL
PB140912BL - 1	PBW	16	0.0126	-	7470A	PPB	19 Nov 2021 12:10:41	U	SMPL
PB140912BS - 1	LCSW	14401	3.3879	-	7470A	PPB	19 Nov 2021 12:12:58	U	SMPL
M4756-03 - 1	DRUM-1-2	-24	0.0032	-	7470A	PPB	19 Nov 2021 12:15:15	U	SMPL
M4756-03DUP - 1	DRUM-1-2DUP	12	0.0117	-	7470A	PPB	19 Nov 2021 12:17:31	U	SMPL
CCV88 - 1	CCV88	21070	4.9527	-	7470A	PPB	19 Nov 2021 12:19:46	U	SMPL
CCB88 - 1	CCB88	-51	-0.0031	-	7470A	PPB	19 Nov 2021 12:22:02	U	SMPL
M4756-03MS - 1	DRUM-1-2MS	14290	3.3618	-	7470A	PPB	19 Nov 2021 12:24:18	U	SMPL
M4756-03MSD - 1	DRUM-1-2MSD	14536	3.4196	-	7470A	PPB	19 Nov 2021 12:26:33	U	SMPL
PB140913BL - 1	PBW	85	0.0288	-	7470A	PPB	19 Nov 2021 12:31:03	U	SMPL
PB140913BS - 1	LCSW	19980	4.6969	-	7470A	PPB	19 Nov 2021 12:33:19	U	SMPL
M4754-04 - 1	TP-12-COMP	54	0.0215	-	7470A	PPB	19 Nov 2021 12:47:30	U	SMPL
M4765-03 - 1	WC-2	45	0.0194	-	7470A	PPB	19 Nov 2021 12:49:46	U	SMPL
M4767-03 - 1	TP-8-COMP	67	0.0246	-	7470A	PPB	19 Nov 2021 12:52:02	U	SMPL
M4767-06 - 1	TP-13-COMP	125	0.0382	-	7470A	PPB	19 Nov 2021 12:54:19	U	SMPL
M4767-06DUP - 1	TP-13-COMP DUP	54	0.0215	-	7470A	PPB	19 Nov 2021 12:56:36	U	SMPL
M4767-06MS - 1	TP-13-COMPMS	17454	4.1042	-	7470A	PPB	19 Nov 2021 12:58:52	U	SMPL
CCV89 - 1	CCV89	21083	4.9557	-	7470A	PPB	19 Nov 2021 13:01:07	U	SMPL
CCB89 - 1	CCB89	-55	-0.0041	-	7470A	PPB	19 Nov 2021 13:03:22	U	SMPL
M4767-06MSD - 1	TP-13-COMPMSD	18594	4.3717	-	7470A	PPB	19 Nov 2021 13:05:38	U	SMPL
PB140914BL - 1	PBW	55	0.0217	-	7470A	PPB	19 Nov 2021 13:10:18	U	SMPL
PB140914BS - 1	LCSW	16046	3.7739	-	7470A	PPB	19 Nov 2021 13:12:33	U	SMPL
M4746-01 - 1	MW-1	7064	1.6663	-	7470A	PPB	19 Nov 2021 13:14:48	U	SMPL
M4746-01DUP - 1	MW-1DUP	6296	1.4861	-	7470A	PPB	19 Nov 2021 13:17:04	U	SMPL
M4746-01MS - 1	MW-1MS	21002	4.9367	-	7470A	PPB	19 Nov 2021 13:22:16	U	SMPL
M4746-01MSD - 1	MW-1MSD	21011	4.9388	-	7470A	PPB	19 Nov 2021 13:24:32	U	SMPL
PB140887TB - 1	PB140845TB	-58	-0.0048	-	7470A	PPB	19 Nov 2021 13:26:49	U	SMPL
PB140912TB - 1	PB140885TB	42	0.0187	-	7470A	PPB	19 Nov 2021 13:29:06	U	SMPL
CCV90 - 1	CCV90	20917	4.9168	-	7470A	PPB	19 Nov 2021 13:33:37	U	SMPL
CCB90 - 1	CCB90	-50	-0.0029	-	7470A	PPB	19 Nov 2021 13:35:52	U	SMPL
M4738-08LX5 - 1		6	0.0102	-	7470A	PPB	19 Nov 2021 13:38:08	U	SMPL
M4738-08A - 1		17525	4.1209	-	7470A	PPB	19 Nov 2021 13:40:23	U	SMPL
M4756-03LX5 - 1		-30	0.0018	-	7470A	PPB	19 Nov 2021 13:42:39	U	SMPL
M4756-03A - 1		17460	4.1056	-	7470A	PPB	19 Nov 2021 13:44:54	U	SMPL
M4767-06LX5 - 1		14	0.0121	-	7470A	PPB	19 Nov 2021 13:47:11	U	SMPL
M4767-06A - 1		17504	4.1160	-	7470A	PPB	19 Nov 2021 13:49:27	U	SMPL
M4746-01LX5 - 1		1466	0.3528	-	7470A	PPB	19 Nov 2021 13:51:43	U	SMPL
M4746-01A - 1		22067	5.1866	-	7470A	PPB	19 Nov 2021 13:54:00	U	SMPL
PB140913TB - 1	PB140888TB	-41	-0.0008	-	7470A	PPB	19 Nov 2021 13:56:17	U	SMPL
CCV91 - 1	CCV91	21130	4.9668	-	7470A	PPB	19 Nov 2021 13:58:33	U	SMPL
CCB91 - 1	CCB91	-66	-0.0066	-	7470A	PPB	19 Nov 2021 14:00:49	U	SMPL