



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/22/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 11/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 11/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117466

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4754-01	TP-4-COMP	1	1.15	8.81	9.96	9.13	90.6	
M4754-05	TP-4-COMP	2	1.17	8.43	9.6	9.23	95.6	
M4769-01	01-A-01-B-01-C	3	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M4769-02	02-A-02-B-02-C	4	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M4769-03	03-A-03-B-03-C	5	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M4769-04	04-A-04-B-04-C	6	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M4769-05	05-A-05-B-05-C	7	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M4769-06	06-A-06-B-06-C	8	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M4769-07	07-A-07-B-07-C	9	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M4770-01	CC-1	10	1.14	8.40	9.54	8.96	93.1	
M4770-03	CC-2	11	1.18	8.35	9.53	8.98	93.4	
M4770-05	CC-3	12	1.18	8.42	9.6	8.53	87.3	
M4770-07	CC-4	13	1.19	8.56	9.75	9.22	93.8	
M4776-01	RO-72-11948	14	1.18	8.79	9.97	6.38	59.2	GEL MATRIX
M4777-01	VNJ-223	15	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
M4781-01	ETGI-354	16	1.14	8.82	9.96	8.2	80.0	
M4782-01	CL-02-111921	17	1.14	8.79	9.93	8.86	87.8	
M4782-02	CL-02-111921-EPH-2	18	1.16	8.76	9.92	8.84	87.7	
M4783-01	RBR-200033	19	1.16	8.81	9.97	8.62	84.7	
M4783-02	RBR-200033-EPH-2	20	1.16	8.81	9.97	8.62	84.7	
M4784-05	SVOC-GPC-BLANK	21	1.00	1.00	2.00	2.00	100.0	
M4784-06	PEST-GPC-BLANK	22	1.00	1.00	2.00	2.00	100.0	
M4784-07	PEST-GPC-BLANK-SPIKE	23	1.00	1.00	2.00	2.00	100.0	
M4784-08	PCB-GPC-BLANK	24	1.00	1.00	2.00	2.00	100.0	
M4784-09	PCB-GPC-BLANK-SPIKE	25	1.00	1.00	2.00	2.00	100.0	
M4784-10	SVOC-GPC2-BLANK	26	1.00	1.00	2.00	2.00	100.0	
M4784-11	PEST-GPC2-BLANK	27	1.00	1.00	2.00	2.00	100.0	
M4784-12	PEST-GPC2-BLANK-SPIKE	28	1.00	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/22/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 11/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 11/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117466

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4784-13	PCB-GCP2-BLANK	29	1.00	1.00	2.00	2.00	100.0	
M4784-14	PCB-GCP2-BLANK-SPIKE	30	1.00	1.00	2.00	2.00	100.0	
M4786-01	RO-279	31	1.14	8.37	9.51	5.44	51.4	GEL MATRIX
M4786-02	RO-279	32	1.17	8.51	9.68	5.27	48.2	GEL MATRIX
M4786-03	RO-357	33	1.19	8.37	9.56	7.62	76.8	GEL MATRIX
M4786-04	RO-357	34	1.13	8.71	9.84	8.7	86.9	
M4789-01	WC-A1-COMP-QAQC-A	35	1.18	8.36	9.54	8.84	91.6	
M4790-01	WC-C1-COMP-QAQC-A	36	1.19	8.67	9.86	9.07	90.9	
M4791-01	WC-C2-COMP-QAQC-A	37	1.16	8.68	9.84	9.11	91.6	
M4792-01	WC-6-(4.5-8)-QAQC	38	1.15	8.75	9.9	9.19	91.9	
M4793-01	WC-3-(4.5-10)-QAQC	39	1.16	8.44	9.6	8.92	91.9	
M4794-01	WC-9-(4.5-10)-QAQC	40	1.13	8.48	9.61	8.95	92.2	
M4795-01	HAR-22	41	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
M4795-03	HAR-23	42	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
M4795-05	HAR-24	43	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
M4795-07	HAR-25	44	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
M4796-01	EO-03-111921	45	1.13	8.83	9.96	8.65	85.2	
M4796-02	EO-03-111921-EPH-2	46	1.14	8.77	9.91	8.74	86.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$