



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117493

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/22/2021	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	11/22/2021	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	11/22/2021	16:19
4	ICV	1	Water	NA	NA	20.4	7.02	11/22/2021	16:20
5	CCV1	1	Water	NA	NA	20.3	2.01	11/22/2021	16:22
6	M4803-04	1	Water	NA	NA	20.1	4.69	11/22/2021	16:24
7	M4803-05	1	Water	NA	NA	20.1	8.04	11/22/2021	16:25
8	M4804-02	1	Water	NA	NA	20.2	4.56	11/22/2021	16:30
9	M4804-03	1	Water	NA	NA	20.5	7.62	11/22/2021	16:32
10	M4807-04	1	Water	NA	NA	21.4	6.72	11/22/2021	16:35
11	M4809-01	1	Water	NA	NA	20.4	6.83	11/22/2021	16:40
12	M4809-02	1	Water	NA	NA	20.5	6.70	11/22/2021	16:42
13	M4809-03	1	Water	NA	NA	20.7	6.82	11/22/2021	16:45
14	M4809-04	1	Water	NA	NA	20.3	6.35	11/22/2021	16:46
15	M4809-06	1	Water	NA	NA	20.3	6.32	11/22/2021	16:50
16	CCV2	1	Water	NA	NA	20.4	12.02	11/22/2021	17:00
17	M4809-06DUP	1	Water	NA	NA	20.4	6.33	11/22/2021	17:01
18	CCV3	1	Water	NA	NA	20.4	2.01	11/22/2021	17:03

WORKLIST(Hardcopy Internal Chain)

1311493

WorkList Name : ph m4804 WorkList ID : 154583 Department : Wet-Chemistry Date : 11-22-2021 15:24:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/22/2021	Water	M4803-04	pH	Cool 4 deg C	PSEG03	F11	261	11/22/2021	9040C
11/22/2021	Water	M4803-05	pH	Cool 4 deg C	PSEG03	F11	262	11/22/2021	9040C
11/22/2021	Water	M4804-02	pH	Cool 4 deg C	PSEG03	F11	40460	11/22/2021	9040C
11/22/2021	Water	M4804-03	pH	Cool 4 deg C	PSEG03	F11	40444	11/22/2021	9040C
11/22/2021	Water	M4807-04	pH	Cool 4 deg C	PSEG03	F11	TOT-COMPOSITE	11/22/2021	9040C
11/22/2021	Water	M4809-01	pH	Cool 4 deg C	PSEG03	F11	1975	11/22/2021	9040C
11/22/2021	Water	M4809-02	pH	Cool 4 deg C	PSEG03	F11	1976	11/22/2021	9040C
11/22/2021	Water	M4809-03	pH	Cool 4 deg C	PSEG03	F11	1977	11/22/2021	9040C
11/22/2021	Water	M4809-04	pH	Cool 4 deg C	PSEG03	F11	1978	11/22/2021	9040C
11/22/2021	Water	M4809-06	pH	Cool 4 deg C	PSEG03	F11	1994	11/22/2021	9040C

Date/Time 11-22-21 16:10
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: JD (CSM)

Date/Time 11-22-21 18:00
 Raw Sample Received by: JD (CSM)
 Raw Sample Relinquished by: JR (WC)