



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117495

Slope : 97.6

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/22/2021	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	11/22/2021	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	11/22/2021	16:19
4	ICV	1	Water	NA	NA	20.4	7.02	11/22/2021	16:20
5	CCV1	1	Water	NA	NA	20.3	2.01	11/22/2021	16:22
6	M4812-01	1	Solid	20.03	20	20.5	11.42	11/22/2021	17:20
7	M4812-01DUP	1	Solid	20.04	20	20.6	11.43	11/22/2021	17:21
8	CCV2	1	Water	NA	NA	20.8	12.03	11/22/2021	17:25

WORKLIST(Hardcopy Internal Chain)

2B112495
13112495

WorkList Name : PH S M4812 WorkList ID : 154595 Department : Wet-Chemistry Date : 11-22-2021 16:50:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/22/2021	Solid	M4812-01	pH	Cool 4 deg C	WALS01	D11	SETTING-TANK-01	11/22/2021	9045D

Date/Time 11-22-21 16:50
 Raw Sample Received by: JP COPI
 Raw Sample Relinquished by: JD CSM

Date/Time 11-22-21 18:30
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JP COPI