



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117501

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	11/23/2021	09:15
2	M4756-03		1	10	100.00	>212.0	>212.0	11/23/2021	09:40
3	M4756-03DUP		1	10	100.00	>212.0	>212.0	11/23/2021	10:05
4	M4804-02		1	12	100.00	>212.0	>212.0	11/23/2021	10:35
5	M4804-03		1	11	100.00	>212.0	>212.0	11/23/2021	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117501

WorkList Name : fp112221 WorkList ID : 154588 Department : Wet-Chemistry Date : 11-22-2021 16:03:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/23/2021	Water	M4756-03	Flash Point	Cool 4 deg C	PSEG03	D11	DRUM-1-2	11/18/2021	1010B
11/29/2021	Water	M4804-02	Flash Point	Cool 4 deg C	PSEG03	F11	40460	11/22/2021	1010B
11/29/2021	Water	M4804-03	Flash Point	Cool 4 deg C	PSEG03	F11	40444	11/22/2021	1010B
11/29/2021	Water	M4804-05	Flash Point	Cool 4 deg C	PSEG03	F11	20872	11/22/2021	1010B

Date/Time 11/23/21 08:45
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JG (WC)

Date/Time 11/23/21 11:10
 Raw Sample Received by: JG (WC)
 Raw Sample Relinquished by: KS (WC)