

LB117505


7470A

INSTRUMENT ID: CVI

Linear

 μ Abs.:

37292



A= 0.0000e+000

B= 2.6644e-004

C= -5.4346e-002

Rho= 0.9993759

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	358	0.000	358				
0.2	0.200	0.258	0.058	1171	0.0 %	1171				
2.5	2.500	2.277	-0.223	8750	0.0 %	8750				
5.0	5.000	5.077	0.077	19259	0.0 %	19259				
7.5	7.500	7.665	0.165	28973	0.0 %	28973				
10.0	10.000	9.882	-0.118	37292	0.0 %	37292				

0/01

29

-9

2

2

-1

LB117505 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Nov 2021 12:53:59					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	358	-	0.0000	7470A	PPB	23 Nov 2021 13:30:33	S	Std
0.2 - 1	50.2	1171	-	0.2000	7470A	PPB	23 Nov 2021 13:32:50	S	Std
2.5 - 1	52.5	8750	-	2.5000	7470A	PPB	23 Nov 2021 13:35:07	S	Std
5.0 - 1	55	19259	-	5.0000	7470A	PPB	23 Nov 2021 13:37:24	S	Std
7.5 - 1	538	28973	-	7.5000	7470A	PPB	23 Nov 2021 13:39:40	S	Std
10.0 - 1	558	37292	-	10.0000	7470A	PPB	23 Nov 2021 13:46:21	S	Std
ICV90 - 1	ICV90	16366	4.3063	-	7470A	PPB	23 Nov 2021 13:49:22	U	SMPL
ICB90 - 1	ICB90	77	-0.0338	-	7470A	PPB	23 Nov 2021 13:53:46	U	SMPL
CCV04 - 1	CCV04	20090	5.2985	-	7470A	PPB	23 Nov 2021 13:56:00	U	SMPL
CCB04 - 1	CCB04	30	-0.0464	-	7470A	PPB	23 Nov 2021 13:58:16	U	SMPL
CRA - 1	CRA	1021	0.2177	-	7470A	PPB	23 Nov 2021 14:07:32	U	SMPL
HighStd - 1	HighStd	38168	10.1153	-	7470A	PPB	23 Nov 2021 14:09:47	U	SMPL
ChkStd - 1	ChkStd	25780	6.8146	-	7470A	PPB	23 Nov 2021 14:12:02	U	SMPL
PB140958BL - 1	PBW	114	-0.0240	-	7470A	PPB	23 Nov 2021 14:16:51	U	SMPL
PB140958BS - 1	LCSW	16610	4.3713	-	7470A	PPB	23 Nov 2021 14:19:07	U	SMPL
M4754-06 - 1	TP-4-COMP	414	0.0560	-	7470A	PPB	23 Nov 2021 14:21:23	U	SMPL
M4754-06DUP - 1	TP-4-COMPDUP	473	0.0717	-	7470A	PPB	23 Nov 2021 14:23:39	U	SMPL
M4754-06MS - 1	TP-4-COMPMS	15338	4.0324	-	7470A	PPB	23 Nov 2021 14:25:56	U	SMPL
M4754-06MSD - 1	TP-4-COMPMSD	16135	4.2447	-	7470A	PPB	23 Nov 2021 14:28:13	U	SMPL
M4781-02 - 1	ETGI-354	409	0.0546	-	7470A	PPB	23 Nov 2021 14:30:29	U	SMPL
CCV05 - 1	CCV05	20257	5.3430	-	7470A	PPB	23 Nov 2021 14:32:45	U	SMPL
CCB05 - 1	CCB05	5	-0.0530	-	7470A	PPB	23 Nov 2021 14:35:00	U	SMPL
M4787-01 - 1	T-1	475	0.0722	-	7470A	PPB	23 Nov 2021 14:37:15	U	SMPL
PB140960BL - 1	PBW	436	0.0618	-	7470A	PPB	23 Nov 2021 14:39:30	U	SMPL
PB140960BS - 1	LCSW	15281	4.0172	-	7470A	PPB	23 Nov 2021 14:41:45	U	SMPL
M4737-04 - 1	COMP-1A-20910-15	934	0.1945	-	7470A	PPB	23 Nov 2021 14:44:01	U	SMPL
M4737-04DUP - 1	COMP-1A-20910-15DUP	949	0.1985	-	7470A	PPB	23 Nov 2021 14:46:17	U	SMPL
M4737-04MS - 1	COMP-1A-20910-15MS	14252	3.7430	-	7470A	PPB	23 Nov 2021 14:48:33	U	SMPL
M4737-04MSD - 1	COMP-1A-20910-15MSD	14481	3.8040	-	7470A	PPB	23 Nov 2021 14:50:49	U	SMPL
M4737-05 - 1	COMP-2A-20916-21	1047	0.2246	-	7470A	PPB	23 Nov 2021 14:53:06	U	SMPL
M4737-06 - 1	COMP-3A-20922-24	985	0.2081	-	7470A	PPB	23 Nov 2021 14:55:23	U	SMPL
PB140958TB - 1	PB140932TB	458	0.0677	-	7470A	PPB	23 Nov 2021 14:57:39	U	SMPL
CCV06 - 1	CCV06	20278	5.3486	-	7470A	PPB	23 Nov 2021 14:59:54	U	SMPL
CCB06 - 1	CCB06	39	-0.0440	-	7470A	PPB	23 Nov 2021 15:06:47	U	SMPL
M4754-06LX5 - 1		23	-0.0482	-	7470A	PPB	23 Nov 2021 15:09:03	U	SMPL
M4754-06A - 1		16490	4.3393	-	7470A	PPB	23 Nov 2021 15:11:18	U	SMPL
M4737-04LX5 - 1		133	-0.0189	-	7470A	PPB	23 Nov 2021 15:13:34	U	SMPL
M4737-04A - 1		14440	3.7931	-	7470A	PPB	23 Nov 2021 15:15:50	U	SMPL
CCV07 - 1	CCV07	19870	5.2399	-	7470A	PPB	23 Nov 2021 15:18:05	U	SMPL
CCB07 - 1	CCB07	-28	-0.0618	-	7470A	PPB	23 Nov 2021 15:20:22	U	SMPL