



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117512

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/23/2021	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	11/23/2021	16:01
3	CAL3	1	Water	NA	NA	20.3	10.03	11/23/2021	16:03
4	ICV	1	Water	NA	NA	20.6	6.99	11/23/2021	16:05
5	CCV1	1	Water	NA	NA	20.4	2.01	11/23/2021	16:10
6	M4817-01	1	Water	NA	NA	20.2	7.49	11/23/2021	16:15
7	M4817-01DUP	1	Water	NA	NA	20.2	7.50	11/23/2021	16:16
8	CCV2	1	Water	NA	NA	20.3	12.03	11/23/2021	16:20

WORKLIST(Hardcopy Internal Chain)

VB 11F512

WorkList Name : ph m4817 WorkList ID : 154625 Department : Wet-Chemistry Date : 11-23-2021 15:43:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/23/2021	Water	M4817-01	pH	Cool 4 deg C	PSEG03	F12	A6879	11/23/2021	9040C

Date/Time 11/23/21 15:43
 Raw Sample Received by: JP 1609
 Raw Sample Relinquished by: J.C. (su)

Date/Time 11/23/21 17:10
 Raw Sample Received by: J.C. (su)
 Raw Sample Relinquished by: JP 1609