

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/24/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 11/23/2021 16:15

DATE & TIME OUT1: 11/24/2021 08:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 103 °C

DATE & TIME IN2: 11/24/2021 08:30

DATE & TIME OUT2: 11/24/2021 10:00

Run Number: LB117513

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117513BL	LB117513BL	93.6417	93.6417	100	93.6417	93.6417	93.6417	0.0000	0
2	M4817-01	A6879	89.1566	89.1566	100	89.3246	89.3246	89.3246	0.1680	1680
3	M4817-01DUP	A6879DUP	75.5465	75.5465	100	75.7191	75.7191	75.7191	0.1726	1726

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

117513

WorkList Name : ts m4817 w WorkList ID : 154623 Department : Wet-Chemistry Date : 11-23-2021 15:42:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/30/2021	Water	M4817-01	TS	Cool 4 deg C	PSEG03	F12	A6879	11/23/2021	SM2540 B

Date/Time 11-23-21 15:45
 Raw Sample Received by: JP
 Raw Sample Relinquished by: JP

Date/Time 11-23-21
 Raw Sample Received by: JP
 Raw Sample Relinquished by: JP