



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/24/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 11/23/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 11/24/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117515

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4804-08	20872	1	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
M4818-01	TR-04-112321	2	1.14	8.80	9.94	9.22	91.8	
M4818-02	TR-04-112321-EPH-2	3	1.19	8.66	9.85	9.26	93.2	
M4819-01	OR-02-112321	4	1.18	8.42	9.6	8.96	92.4	
M4819-02	OR-02-112321-EPH-2	5	1.19	8.57	9.76	9.1	92.3	
M4822-01	MASTIC	6	1.00	1.00	2.00	2.00	100.0	MASTIC
M4823-01	FDA6IV-END-1-1	7	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-02	FDA6IV-END-1-2	8	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-03	SEC-2021-48A	9	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-04	SEC-2021-48B	10	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-05	SEC-2021-49A	11	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-06	SEC-2021-49B	12	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-07	SEC-2021-50A	13	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-08	SEC-2021-50B	14	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-09	SEC-2021-51A	15	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-10	SEC-2021-51B	16	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-11	SEC-2021-52A	17	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4823-12	SEC-2021-52B	18	1.00	1.00	2.00	2.00	100.0	PILC REEL
M4824-01	T-1	19	1.17	8.58	9.75	8.65	87.2	
M4824-02	T-2	20	1.19	8.73	9.92	8.75	86.6	
M4824-03	T-3	21	1.18	8.79	9.97	8.83	87.0	
M4824-04	T-4	22	1.15	8.39	9.54	8.32	85.5	
M4824-04D	T-4DUP	23	1.16	8.40	9.56	8.35	85.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$