

LB117522


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

36777



A= 0.0000e+000

B= 2.7022e-004

C= -9.0897e-003

Rho= 0.9997699

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.002	-0.002	27	0.000	27					0/00
0.2	0.200	0.210	0.010	809	0.0 %	809					5
2.5	2.500	2.391	-0.109	8882	0.0 %	8882					-4
5.0	5.000	5.139	0.139	19053	0.0 %	19053					3
7.5	7.500	7.533	0.033	27912	0.0 %	27912					0
10.0	10.000	9.929	-0.071	36777	0.0 %	36777					-1

LB117522 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Nov 2021 09:21:31

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	27	-	0.00007470A	PPB	24 Nov 2021 09:31:14	S	Std
0.2 - 1	50.2	809	-	0.20007470A	PPB	24 Nov 2021 09:33:31	S	Std
2.5 - 1	52.5	8882	-	2.50007470A	PPB	24 Nov 2021 09:35:48	S	Std
5.0 - 1	55	19053	-	5.00007470A	PPB	24 Nov 2021 09:40:50	S	Std
7.5 - 1	57.5	27912	-	7.50007470A	PPB	24 Nov 2021 09:43:06	S	Std
10.0 - 1	510	36777	-	10.00007470A	PPB	24 Nov 2021 09:47:56	S	Std
ICV91 - 1	ICV91	15340	4.1360	-7470A	PPB	24 Nov 2021 09:50:35	U	SMPL
ICB91 - 1	ICB91	-63	-0.0261	-7470A	PPB	24 Nov 2021 09:52:50	U	SMPL
CCV08 - 1	CCV08	18956	5.1132	-7470A	PPB	24 Nov 2021 09:55:05	U	SMPL
CCB08 - 1	CCB08	-58	-0.0248	-7470A	PPB	24 Nov 2021 09:57:20	U	SMPL
CRA - 1	CRA	797	0.2063	-7470A	PPB	24 Nov 2021 09:59:35	U	SMPL
HighStd - 1	HighStd	36574	9.8738	-7470A	PPB	24 Nov 2021 10:01:51	U	SMPL
ChkStd - 1	ChkStd	25991	7.0141	-7470A	PPB	24 Nov 2021 10:04:07	U	SMPL
PB140986BL - 1	PBW	-98	-0.0356	-7470A	PPB	24 Nov 2021 10:06:23	U	SMPL
PB140986BS - 1	LCSW	15270	4.1171	-7470A	PPB	24 Nov 2021 10:08:40	U	SMPL
M4803-01 - 1	266	58	0.0066	-7470A	PPB	24 Nov 2021 10:10:56	U	SMPL
M4803-02 - 1	260	11	-0.0061	-7470A	PPB	24 Nov 2021 10:13:13	U	SMPL
M4803-03 - 1	263	16	-0.0048	-7470A	PPB	24 Nov 2021 10:15:29	U	SMPL
M4812-02 - 1	SETTING-TANK-01	12	-0.0058	-7470A	PPB	24 Nov 2021 10:17:47	U	SMPL
M4812-02DUP - 1	SETTING-TANK-01DUP	26	-0.0021	-7470A	PPB	24 Nov 2021 10:20:02	U	SMPL
CCV09 - 1	CCV09	18745	5.0561	-7470A	PPB	24 Nov 2021 10:22:18	U	SMPL
CCB09 - 1	CCB09	-59	-0.0250	-7470A	PPB	24 Nov 2021 10:24:34	U	SMPL
M4812-02MS - 1	SETTING-TANK-01MS	14693	3.9612	-7470A	PPB	24 Nov 2021 10:26:49	U	SMPL
M4812-02MSD - 1	SETTING-TANK-01MSD	13801	3.7202	-7470A	PPB	24 Nov 2021 10:29:04	U	SMPL
PB140987BL - 1	PBW	-51	-0.0229	-7470A	PPB	24 Nov 2021 10:31:19	U	SMPL
PB140987BS - 1	LCSW	14384	3.8777	-7470A	PPB	24 Nov 2021 10:33:35	U	SMPL
M4803-06 - 1	261-262	-38	-0.0194	-7470A	PPB	24 Nov 2021 10:35:50	U	SMPL
M4809-05 - 1	75-78	8	-0.0069	-7470A	PPB	24 Nov 2021 10:38:07	U	SMPL
M4809-05DUP - 1	75-78DUP	12	-0.0058	-7470A	PPB	24 Nov 2021 10:40:23	U	SMPL
M4809-05MS - 1	75-78MS	16477	4.4433	-7470A	PPB	24 Nov 2021 10:42:40	U	SMPL
M4809-05MSD - 1	75-78MSD	14706	3.9647	-7470A	PPB	24 Nov 2021 10:44:57	U	SMPL
M4809-06 - 1	1994	-57	-0.0245	-7470A	PPB	24 Nov 2021 10:47:13	U	SMPL
CCV10 - 1	CCV10	18746	5.0564	-7470A	PPB	24 Nov 2021 10:49:28	U	SMPL
CCB10 - 1	CCB10	-53	-0.0234	-7470A	PPB	24 Nov 2021 10:51:44	U	SMPL
PB140988BL - 1	PBW	5	-0.0077	-7470A	PPB	24 Nov 2021 10:53:59	U	SMPL
PB140988BS - 1	LCSW	14064	3.7913	-7470A	PPB	24 Nov 2021 10:56:15	U	SMPL
M4807-04 - 1	TOT-COMPOSITE	1995	0.5300	-7470A	PPB	24 Nov 2021 10:58:31	U	SMPL
M4807-04DUP - 1	TOT-COMPOSITEDUP	2066	0.5492	-7470A	PPB	24 Nov 2021 11:00:47	U	SMPL
M4807-04MS - 1	TOT-COMPOSITEMS	13175	3.5510	-7470A	PPB	24 Nov 2021 11:06:09	U	SMPL
M4807-04MSD - 1	TOT-COMPOSITEMSD	14049	3.7872	-7470A	PPB	24 Nov 2021 11:08:25	U	SMPL
PB140986TB - 1	PB140963TB	-43	-0.0207	-7470A	PPB	24 Nov 2021 11:10:42	U	SMPL
PB140987TB - 1	PB140973TB	22	-0.0031	-7470A	PPB	24 Nov 2021 11:12:58	U	SMPL
CCV11 - 1	CCV11	17148	4.6246	-7470A	PPB	24 Nov 2021 11:15:15	U	SMPL
CCB11 - 1	CCB11	-24	-0.0156	-7470A	PPB	24 Nov 2021 11:17:31	U	SMPL
M4812-02LX5 - 1		21	-0.0034	-7470A	PPB	24 Nov 2021 11:19:47	U	SMPL
M4812-02A - 1		14270	3.8469	-7470A	PPB	24 Nov 2021 11:22:03	U	SMPL
M4809-05LX5 - 1		-38	-0.0194	-7470A	PPB	24 Nov 2021 11:24:18	U	SMPL
M4809-05A - 1		14405	3.8834	-7470A	PPB	24 Nov 2021 11:26:34	U	SMPL
M4807-04LX5 - 1		422	0.1049	-7470A	PPB	24 Nov 2021 11:28:49	U	SMPL
M4807-04A - 1		13920	3.7523	-7470A	PPB	24 Nov 2021 11:31:05	U	SMPL
CCV12 - 1	CCV12	17100	4.6116	-7470A	PPB	24 Nov 2021 11:33:21	U	SMPL
CCB12 - 1	CCB12	-57	-0.0245	-7470A	PPB	24 Nov 2021 11:35:38	U	SMPL