

LB117526


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38725



A= 0.0000e+000

B= 2.5498e-004 *slope*C= 7.0313e-003 *y-intercept*

Rho= 0.9995883

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.019	0.019	48	0.000	48					—
0.05	0.050					205	0				13
0.20	0.200	0.226	0.026	857	0.0 %	857					-6
2.50	2.500	2.344	-0.156	9166	0.0 %	9166					2
5.00	5.000	5.087	0.087	19924	0.0 %	19924					2
7.50	7.500	7.642	0.142	29944	0.0 %	29944					-1
10.0	10.000	9.881	-0.119	38725	0.0 %	38725					

LB117526
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	48	-		0 SFAM01.1	PPB	11/24/2021 12:30	Std
	0.2 S01	857	-		0.2 SFAM01.1	PPB	11/24/2021 12:33	Std
	2.5 S02	9166	-		2.5 SFAM01.1	PPB	11/24/2021 12:35	Std
	5 S03	19924	-		5 SFAM01.1	PPB	11/24/2021 12:37	Std
	7.5 S04	29944	-		7.5 SFAM01.1	PPB	11/24/2021 12:40	Std
	10 S05	38725	-		10 SFAM01.1	PPB	11/24/2021 12:42	Std
ICV090	ICV090	16266	4.1546 -		SFAM01.1	PPB	11/24/2021 12:45	SMPL
ICB090	ICB090	3	0.0078 -		SFAM01.1	PPB	11/24/2021 12:47	SMPL
CCV013	CCV013	19795	5.0545 -		SFAM01.1	PPB	11/24/2021 12:49	SMPL
CCB013	CCB013	-1	0.0068 -		SFAM01.1	PPB	11/24/2021 12:51	SMPL
PB141017BL	PBS017	37	0.0165 -		SFAM01.1	PPB	11/24/2021 12:54	SMPL
M4629-01	GB8K4	128	0.0397 -		SFAM01.1	PPB	11/24/2021 12:56	SMPL
M4629-02	GB8K4D	189	0.0552 -		SFAM01.1	PPB	11/24/2021 12:58	SMPL
M4629-03	GB8K4S	11111	2.8402 -		SFAM01.1	PPB	11/24/2021 13:00	SMPL
M4629-04	GB8K5	517	0.1389 -		SFAM01.1	PPB	11/24/2021 13:03	SMPL
M4629-05	GB8K6	547	0.1465 -		SFAM01.1	PPB	11/24/2021 13:05	SMPL
M4629-06	GB8K7	586	0.1565 -		SFAM01.1	PPB	11/24/2021 13:07	SMPL
M4629-07	GB8K8	560	0.1498 -		SFAM01.1	PPB	11/24/2021 13:10	SMPL
M4629-08	GB8K9	734	0.1942 -		SFAM01.1	PPB	11/24/2021 13:12	SMPL
M4629-09	GB8L0	393	0.1072 -		SFAM01.1	PPB	11/24/2021 13:14	SMPL
M4629-10	GB8L1	531	0.1424 -		SFAM01.1	PPB	11/24/2021 13:16	SMPL
M4629-11	GB8L2	533	0.1429 -		SFAM01.1	PPB	11/24/2021 13:19	SMPL
M4629-12	GB8L3	661	0.1756 -		SFAM01.1	PPB	11/24/2021 13:21	SMPL
M4629-13	GB8L4	590	0.1575 -		SFAM01.1	PPB	11/24/2021 13:23	SMPL
M4629-14	GB8L5	1918	0.4961 -		SFAM01.1	PPB	11/24/2021 13:25	SMPL
M4629-15	GB8L6	946	0.2482 -		SFAM01.1	PPB	11/24/2021 13:28	SMPL
M4629-16	GB8L7	889	0.2337 -		SFAM01.1	PPB	11/24/2021 13:30	SMPL
M4629-17	GB8L8	765	0.2021 -		SFAM01.1	PPB	11/24/2021 13:32	SMPL
M4629-18	GB8L9	3465	0.8906 -		SFAM01.1	PPB	11/24/2021 13:34	SMPL
M4629-19	GB8M0	4822	1.2366 -		SFAM01.1	PPB	11/24/2021 13:37	SMPL
CCV014	CCV014	20267	5.1748 -		SFAM01.1	PPB	11/24/2021 13:39	SMPL
CCB014	CCB014	8	0.0091 -		SFAM01.1	PPB	11/24/2021 13:41	SMPL
M4629-20	GB8M1	234	0.0667 -		SFAM01.1	PPB	11/24/2021 13:44	SMPL
M4629-21	GB8M2	235	0.067 -		SFAM01.1	PPB	11/24/2021 13:46	SMPL
M4629-22	GB8M3	700	0.1855 -		SFAM01.1	PPB	11/24/2021 13:48	SMPL
PB141018BL	PBS018	18	0.0116 -		SFAM01.1	PPB	11/24/2021 13:50	SMPL
M4631-01	GB8M4	3586	0.9214 -		SFAM01.1	PPB	11/24/2021 13:57	SMPL
M4631-02	GB8M4D	3548	0.9117 -		SFAM01.1	PPB	11/24/2021 14:00	SMPL
M4631-03	GB8M4S	13205	3.3741 -		SFAM01.1	PPB	11/24/2021 14:02	SMPL
M4631-04	GB8M5	471	0.1271 -		SFAM01.1	PPB	11/24/2021 14:04	SMPL
M4631-05	GB8M6	2661	0.6855 -		SFAM01.1	PPB	11/24/2021 14:06	SMPL
M4631-06	GB8M7	503	0.1353 -		SFAM01.1	PPB	11/24/2021 14:09	SMPL
M4631-07	GB8M8	561	0.1501 -		SFAM01.1	PPB	11/24/2021 14:11	SMPL
M4631-08	GB8M9	495	0.1332 -		SFAM01.1	PPB	11/24/2021 14:13	SMPL
M4631-09	GB8N0	645	0.1715 -		SFAM01.1	PPB	11/24/2021 14:16	SMPL
M4631-10	GB8N1	870	0.2289 -		SFAM01.1	PPB	11/24/2021 14:18	SMPL

LB117526
INSTRUMENT ID : CV1

M4631-11	GB8N2	761	0.2011 -	SFAM01.1	PPB	11/24/2021 14:20 SMPL
M4631-12	GB8N3	742	0.1962 -	SFAM01.1	PPB	11/24/2021 14:22 SMPL
M4631-13	GB8N4	825	0.2174 -	SFAM01.1	PPB	11/24/2021 14:25 SMPL
M4631-14	GB8N5	203	0.0588 -	SFAM01.1	PPB	11/24/2021 14:27 SMPL
M4631-15	GB8N6	614	0.1636 -	SFAM01.1	PPB	11/24/2021 14:29 SMPL
M4631-16	GB8N7	729	0.1929 -	SFAM01.1	PPB	11/24/2021 14:31 SMPL
CCV015	CCV015	20341	5.1937 -	SFAM01.1	PPB	11/24/2021 14:34 SMPL
CCB015	CCB015	15	0.0109 -	SFAM01.1	PPB	11/24/2021 14:36 SMPL
M4631-17	GB8N8	867	0.2281 -	SFAM01.1	PPB	11/24/2021 14:38 SMPL
M4631-18	GB8N9	1042	0.2727 -	SFAM01.1	PPB	11/24/2021 14:40 SMPL
M4631-19	GB8P0	930	0.2442 -	SFAM01.1	PPB	11/24/2021 14:43 SMPL
M4631-20	GB8P1	590	0.1575 -	SFAM01.1	PPB	11/24/2021 14:45 SMPL
M4631-21	GB8P2	580	0.1549 -	SFAM01.1	PPB	11/24/2021 14:47 SMPL
M4631-22	GB8P3	1329	0.3459 -	SFAM01.1	PPB	11/24/2021 14:49 SMPL
PB141019BL	PBS019	36	0.0162 -	SFAM01.1	PPB	11/24/2021 14:52 SMPL
M4632-01	GB8P4	3253	0.8365 -	SFAM01.1	PPB	11/24/2021 14:54 SMPL
M4632-02	GB8P4D	3930	1.0091 -	SFAM01.1	PPB	11/24/2021 14:56 SMPL
M4632-03	GB8P4S	13930	3.559 -	SFAM01.1	PPB	11/24/2021 15:01 SMPL
M4632-04	GB8P5	238	0.0677 -	SFAM01.1	PPB	11/24/2021 15:04 SMPL
M4632-05	GB8P6	387	0.1057 -	SFAM01.1	PPB	11/24/2021 15:06 SMPL
M4632-06	GB8P7	841	0.2215 -	SFAM01.1	PPB	11/24/2021 15:08 SMPL
M4632-07	GB8P8	924	0.2426 -	SFAM01.1	PPB	11/24/2021 15:11 SMPL
M4632-08	GB8P9	803	0.2118 -	SFAM01.1	PPB	11/24/2021 15:13 SMPL
M4632-09	GB8Q0	721	0.1909 -	SFAM01.1	PPB	11/24/2021 15:15 SMPL
M4632-10	GB8Q1	874	0.2299 -	SFAM01.1	PPB	11/24/2021 15:17 SMPL
M4632-11	GB8Q2	690	0.183 -	SFAM01.1	PPB	11/24/2021 15:20 SMPL
M4632-12	GB8Q3	959	0.2516 -	SFAM01.1	PPB	11/24/2021 15:22 SMPL
M4632-13	GB8Q4	811	0.2138 -	SFAM01.1	PPB	11/24/2021 15:24 SMPL
CCV016	CCV016	21075	5.3808 -	SFAM01.1	PPB	11/24/2021 15:26 SMPL
CCB016	CCB016	3	0.0078 -	SFAM01.1	PPB	11/24/2021 15:29 SMPL
M4632-14	GB8Q5	876	0.2304 -	SFAM01.1	PPB	11/24/2021 15:31 SMPL
M4632-15	GB8Q6	259	0.0731 -	SFAM01.1	PPB	11/24/2021 15:33 SMPL
M4632-16	GB8Q7	828	0.2182 -	SFAM01.1	PPB	11/24/2021 15:36 SMPL
M4632-17	GB8Q8	725	0.1919 -	SFAM01.1	PPB	11/24/2021 15:38 SMPL
M4632-18	GB8Q9	664	0.1763 -	SFAM01.1	PPB	11/24/2021 15:40 SMPL
M4632-19	GB8R0	705	0.1868 -	SFAM01.1	PPB	11/24/2021 15:42 SMPL
M4632-20	GB8R1	404	0.11 -	SFAM01.1	PPB	11/24/2021 15:45 SMPL
M4632-21	GB8R2	735	0.1944 -	SFAM01.1	PPB	11/24/2021 15:47 SMPL
M4632-22	GB8R3	752	0.1988 -	SFAM01.1	PPB	11/24/2021 15:49 SMPL
PB141020BL	PBS020	13	0.0103 -	SFAM01.1	PPB	11/24/2021 15:51 SMPL
M4633-01	GB8R4	777	0.2052 -	SFAM01.1	PPB	11/24/2021 15:54 SMPL
M4633-02	GB8R5	989	0.2592 -	SFAM01.1	PPB	11/24/2021 15:56 SMPL
M4633-03	GB8R6	845	0.2225 -	SFAM01.1	PPB	11/24/2021 15:58 SMPL
M4633-04	GB8R7	817	0.2154 -	SFAM01.1	PPB	11/24/2021 16:01 SMPL
M4633-05	GB8R8	779	0.2057 -	SFAM01.1	PPB	11/24/2021 16:03 SMPL
M4633-06	GB8R9	752	0.1988 -	SFAM01.1	PPB	11/24/2021 16:05 SMPL
M4633-07	GB8S0	969	0.2541 -	SFAM01.1	PPB	11/24/2021 16:07 SMPL

LB117526
INSTRUMENT ID : CV1

M4633-08	GB8S1	768	0.2029 -	SFAM01.1	PPB	11/24/2021 16:10 SMPL
M4633-09	GB8S2	415	0.1128 -	SFAM01.1	PPB	11/24/2021 16:12 SMPL
M4633-10	GB8S3	342	0.0942 -	SFAM01.1	PPB	11/24/2021 16:14 SMPL
CCV017	CCV017	21509	5.4915 -	SFAM01.1	PPB	11/24/2021 16:16 SMPL
CCB017	CCB017	8	0.0091 -	SFAM01.1	PPB	11/24/2021 16:19 SMPL
M4633-11	GB8S4	249	0.0705 -	SFAM01.1	PPB	11/24/2021 16:21 SMPL
M4633-12	GB8S5	717	0.1899 -	SFAM01.1	PPB	11/24/2021 16:23 SMPL
M4633-13	GB8S6	964	0.2528 -	SFAM01.1	PPB	11/24/2021 16:26 SMPL
M4633-14	GB8S6D	931	0.2444 -	SFAM01.1	PPB	11/24/2021 16:28 SMPL
M4633-15	GB8S6S	11814	3.0194 -	SFAM01.1	PPB	11/24/2021 16:33 SMPL
M4633-16	GB8S7	770	0.2034 -	SFAM01.1	PPB	11/24/2021 16:35 SMPL
M4633-17	GB8S8	604	0.161 -	SFAM01.1	PPB	11/24/2021 16:38 SMPL
M4633-18	GB8S9	570	0.1524 -	SFAM01.1	PPB	11/24/2021 16:40 SMPL
M4633-19	GB8T0	581	0.1552 -	SFAM01.1	PPB	11/24/2021 16:42 SMPL
M4633-20	GB8T1	437	0.1185 -	SFAM01.1	PPB	11/24/2021 16:44 SMPL
M4633-21	GB8T2	459	0.1241 -	SFAM01.1	PPB	11/24/2021 16:47 SMPL
M4633-22	GB8T3	408	0.1111 -	SFAM01.1	PPB	11/24/2021 16:49 SMPL
CCV018	CCV018	20904	5.3372 -	SFAM01.1	PPB	11/24/2021 16:51 SMPL
CCB018	CCB018	6	0.0086 -	SFAM01.1	PPB	11/24/2021 16:53 SMPL