



Analytical Summary Report

Analysis Method: 1010A
Parameter: Flash Point
Run Number: LB117532
Thermometer ID: flashpoint

Reviewed By: ketankumar
Supervisor Review By: apatel
Ambient Barometric Pressure (mmHg): 765.00
Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV) | W2743 |

| Seq | LabID | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1 | ICV | 81 | 1 | 7 | 27.00 | 80.6 | 80.3 | 11/24/2021 | 13:25 |
| 2 | M4804-08 | | 1 | 10 | 58.00 | 136.4 | 136.1 | 11/24/2021 | 13:45 |
| 3 | M4804-08DUP | | 1 | 12 | 60.00 | 140.0 | 139.7 | 11/24/2021 | 14:10 |

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

LB 117532

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp112421-1 WorkList ID : 154646 Department : Wet-Chemistry Date : 11-24-2021 09:02:51

| Due Date | Matrix | Sample | Test | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 11/29/2021 | Solid | M4804-08 | Flash Point | Cool 4 deg C | PSEG03 | F11 | 20872 | 11/22/2021 | 1010A |

Date/Time 11/24/21 13:05
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJCEA-lab

Date/Time 11/24/21 14:15
 Raw Sample Received by: RJCEA-lab
 Raw Sample Relinquished by: KS (WC)