

LB117546


7471B

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

51855



A= 0.0000e+000

B= 1.9461e-004

C= -6.0981e-002

Rho= 0.9999400

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.009  | 0.009  | 362   | 0.000      | 362   |       |       |       |       | -  |
| 0.2    | 0.200  | 0.183  | -0.017 | 1254  | 0.0 %      | 1254  |       |       |       |       | -9 |
| 2.5    | 2.500  | 2.502  | 0.002  | 13171 | 0.0 %      | 13171 |       |       |       |       | 0  |
| 5.0    | 5.000  | 5.051  | 0.051  | 26268 | 0.0 %      | 26268 |       |       |       |       | 1  |
| 7.5    | 7.500  | 7.423  | -0.077 | 38458 | 0.0 %      | 38458 |       |       |       |       | -1 |
| 10.0   | 10.000 | 10.031 | 0.031  | 51855 | 0.0 %      | 51855 |       |       |       |       | 0  |

## LB117546 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Nov 2021 09:48:07

| Sample ID       | Extended ID     | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-----------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50              | 362    | -       | 0.0000   | 7471B  | PPB   | 26 Nov 2021 09:54:46 | S    | Std  |
| 0.2 - 1         | 50.2            | 1254   | -       | 0.2000   | 7471B  | PPB   | 26 Nov 2021 09:57:03 | S    | Std  |
| 2.5 - 1         | 52.5            | 13171  | -       | 2.5000   | 7471B  | PPB   | 26 Nov 2021 09:59:20 | S    | Std  |
| 5.0 - 1         | 55              | 26268  | -       | 5.0000   | 7471B  | PPB   | 26 Nov 2021 10:01:37 | S    | Std  |
| 7.5 - 1         | 57.5            | 38458  | -       | 7.5000   | 7471B  | PPB   | 26 Nov 2021 10:03:54 | S    | Std  |
| 10.0 - 1        | 510             | 51855  | -       | 10.0000  | 7471B  | PPB   | 26 Nov 2021 10:08:59 | S    | Std  |
| ICV92 - 1       | ICV92           | 21408  | 4.1053  | -        | 7471B  | PPB   | 26 Nov 2021 10:11:34 | U    | SMPL |
| ICB92 - 1       | ICB92           | 258    | -0.0108 | -        | 7471B  | PPB   | 26 Nov 2021 10:13:50 | U    | SMPL |
| CCV19 - 1       | CCV19           | 27204  | 5.2333  | -        | 7471B  | PPB   | 26 Nov 2021 10:16:05 | U    | SMPL |
| CCB19 - 1       | CCB19           | 231    | -0.0160 | -        | 7471B  | PPB   | 26 Nov 2021 10:18:20 | U    | SMPL |
| CRA - 1         | CRA             | 1257   | 0.1836  | -        | 7471B  | PPB   | 26 Nov 2021 10:20:37 | U    | SMPL |
| HighStd - 1     | HighStd         | 52772  | 10.2091 | -        | 7471B  | PPB   | 26 Nov 2021 10:22:51 | U    | SMPL |
| ChkStd - 1      | ChkStd          | 37042  | 7.1479  | -        | 7471B  | PPB   | 26 Nov 2021 10:25:07 | U    | SMPL |
| PB141028BL - 1  | PB141028BL      | 198    | -0.0224 | -        | 7471B  | PPB   | 26 Nov 2021 10:27:24 | U    | SMPL |
| PB141028BS - 1  | PB141028BS      | 19766  | 3.7857  | -        | 7471B  | PPB   | 26 Nov 2021 10:29:41 | U    | SMPL |
| M4804-01 - 1    | 40442           | 844    | 0.1033  | -        | 7471B  | PPB   | 26 Nov 2021 10:31:57 | U    | SMPL |
| M4807-01 - 1    | 5290            | 1612   | 0.2527  | -        | 7471B  | PPB   | 26 Nov 2021 10:34:13 | U    | SMPL |
| M4807-02 - 1    | 5292            | 674    | 0.0702  | -        | 7471B  | PPB   | 26 Nov 2021 10:36:30 | U    | SMPL |
| M4808-01 - 1    | CHRT-20001      | 597    | 0.0552  | -        | 7471B  | PPB   | 26 Nov 2021 10:38:48 | U    | SMPL |
| M4811-01 - 1    | 20071           | 2310   | 0.3886  | -        | 7471B  | PPB   | 26 Nov 2021 10:41:03 | U    | SMPL |
| CCV20 - 1       | CCV20           | 27539  | 5.2985  | -        | 7471B  | PPB   | 26 Nov 2021 10:43:18 | U    | SMPL |
| CCB20 - 1       | CCB20           | 174    | -0.0271 | -        | 7471B  | PPB   | 26 Nov 2021 10:45:33 | U    | SMPL |
| M4812-01 - 1    | SETTING-TANK-01 | 4304   | 0.7766  | -        | 7471B  | PPB   | 26 Nov 2021 10:47:49 | U    | SMPL |
| M4814-01 - 1    | PL-1-112221     | 2216   | 0.3703  | -        | 7471B  | PPB   | 26 Nov 2021 10:50:05 | U    | SMPL |
| M4818-01 - 1    | TR-04-112321    | 1952   | 0.3189  | -        | 7471B  | PPB   | 26 Nov 2021 10:52:20 | U    | SMPL |
| M4819-01 - 1    | OR-02-112321    | 972    | 0.1282  | -        | 7471B  | PPB   | 26 Nov 2021 10:54:36 | U    | SMPL |
| M4830-08 - 1    | 02-04           | 13498  | 2.5659  | -        | 7471B  | PPB   | 26 Nov 2021 10:56:52 | U    | SMPL |
| M4830-09 - 1    | 05-08           | 21266  | 4.0777  | -        | 7471B  | PPB   | 26 Nov 2021 10:59:08 | U    | SMPL |
| M4837-04 - 1    | 40-42           | 6507   | 1.2054  | -        | 7471B  | PPB   | 26 Nov 2021 11:01:24 | U    | SMPL |
| M4842-01 - 1    | OK-02-112421    | 2868   | 0.4972  | -        | 7471B  | PPB   | 26 Nov 2021 11:03:41 | U    | SMPL |
| M4843-01 - 1    | L-1             | 5468   | 1.0032  | -        | 7471B  | PPB   | 26 Nov 2021 11:05:58 | U    | SMPL |
| CCV21 - 1       | CCV21           | 27564  | 5.3033  | -        | 7471B  | PPB   | 26 Nov 2021 11:08:14 | U    | SMPL |
| CCB21 - 1       | CCB21           | 266    | -0.0092 | -        | 7471B  | PPB   | 26 Nov 2021 11:10:29 | U    | SMPL |
| M4844-01 - 1    | K-1             | 2465   | 0.4187  | -        | 7471B  | PPB   | 26 Nov 2021 11:12:46 | U    | SMPL |
| M4844-01DUP - 1 | K-1DUP          | 2442   | 0.4143  | -        | 7471B  | PPB   | 26 Nov 2021 11:15:02 | U    | SMPL |
| M4844-01MS - 1  | K-1MS           | 22711  | 4.3589  | -        | 7471B  | PPB   | 26 Nov 2021 11:17:17 | U    | SMPL |
| M4844-01MSD - 1 | K-1MSD          | 22157  | 4.2511  | -        | 7471B  | PPB   | 26 Nov 2021 11:19:33 | U    | SMPL |
| M4844-01LX5 - 1 |                 | 713    | 0.0778  | -        | 7471B  | PPB   | 26 Nov 2021 11:21:48 | U    | SMPL |
| M4844-01A - 1   |                 | 23503  | 4.5130  | -        | 7471B  | PPB   | 26 Nov 2021 11:24:04 | U    | SMPL |
| CCV22 - 1       | CCV22           | 27525  | 5.2957  | -        | 7471B  | PPB   | 26 Nov 2021 11:26:20 | U    | SMPL |
| CCB22 - 1       | CCB22           | 284    | -0.0057 | -        | 7471B  | PPB   | 26 Nov 2021 11:28:37 | U    | SMPL |