

LB117549


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35444



A= 0.0000e+000

B= 2.8368e-004

C= -3.9981e-004

Rho= 0.9999435

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.000	-0.000	0	0.000	0					~
0.05	0.050	0.051	0.001	182	0.0 %	182	0				2
0.2	0.200	0.197	-0.003	696	0.0 %	696					-2
2.5	2.500	2.547	0.047	8981	0.0 %	8981					2
5.0	5.000	4.970	-0.030	17522	0.0 %	17522					-1
7.5	7.500	7.430	-0.070	26195	0.0 %	26195					-1
10.0	10.000	10.054	0.054	35444	0.0 %	35444					1

LB117547 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 26 Nov 2021 11:56:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	245.1	PPB	26 Nov 2021 12:08:52	S	Std
0.05 - 1	50.05	182	-	0.0500	245.1	PPB	26 Nov 2021 12:11:09	S	Std
0.2 - 1	50.2	696	-	0.2000	245.1	PPB	26 Nov 2021 12:13:26	S	Std
2.5 - 1	52.5	8981	-	2.5000	245.1	PPB	26 Nov 2021 12:15:43	S	Std
5.0 - 1	55	17522	-	5.0000	245.1	PPB	26 Nov 2021 12:17:59	S	Std
7.5 - 1	57.5	26195	-	7.5000	245.1	PPB	26 Nov 2021 12:20:15	S	Std
10.0 - 1	510	35444	-	10.0000	245.1	PPB	26 Nov 2021 12:22:30	S	Std
ICV93 - 1	ICV93	14272	4.0482	-245.1		PPB	26 Nov 2021 12:29:40	U	SMPL
ICB93 - 1	ICB93	56	0.0155	-245.1		PPB	26 Nov 2021 12:31:55	U	SMPL
CCV23 - 1	CCV23	17593	4.9903	-245.1		PPB	26 Nov 2021 12:34:09	U	SMPL
CCB23 - 1	CCB23	28	0.0075	-245.1		PPB	26 Nov 2021 12:36:24	U	SMPL
CRA - 1	CRA	152	0.0427	-245.1		PPB	26 Nov 2021 12:38:39	U	SMPL
HighStd - 1	HighStd	35819	10.1606	-245.1		PPB	26 Nov 2021 12:40:54	U	SMPL
ChkStd - 1	ChkStd	25648	7.2753	-245.1		PPB	26 Nov 2021 12:43:10	U	SMPL
PB141031BL - 1	PBW	36	0.0098	-245.1		PPB	26 Nov 2021 12:47:59	U	SMPL
PB141031BS - 1	LCSW	13803	3.9152	-245.1		PPB	26 Nov 2021 12:50:15	U	SMPL
M4806-07 - 1	14B-(1-4)-COMP	818	0.2316	-245.1		PPB	26 Nov 2021 12:52:31	U	SMPL
M4806-07DUP - 1	14B-(1-4)-COMPDUP	788	0.2231	-245.1		PPB	26 Nov 2021 12:54:48	U	SMPL
M4806-07MS - 1	14B-(1-4)-COMPMS	16738	4.7478	-245.1		PPB	26 Nov 2021 12:57:05	U	SMPL
M4806-07MSD - 1	14B-(1-4)-COMPMSD	15712	4.4567	-245.1		PPB	26 Nov 2021 13:01:55	U	SMPL
M4820-07 - 1	14B-(1-4)-COMP	668	0.1891	-245.1		PPB	26 Nov 2021 13:04:10	U	SMPL
CCV24 - 1	CCV24	17529	4.9721	-245.1		PPB	26 Nov 2021 13:06:26	U	SMPL
CCB24 - 1	CCB24	106	0.0297	-245.1		PPB	26 Nov 2021 13:08:42	U	SMPL
M4806-07LX5 - 1		139	0.0390	-245.1		PPB	26 Nov 2021 13:10:57	U	SMPL
M4806-07A - 1		14770	4.1895	-245.1		PPB	26 Nov 2021 13:13:13	U	SMPL
CCV25 - 1	CCV25	17722	5.0269	-245.1		PPB	26 Nov 2021 13:18:10	U	SMPL
CCB25 - 1	CCB25	57	0.0158	-245.1		PPB	26 Nov 2021 13:20:25	U	SMPL