

LB117548


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

39160



A= 0.0000e+000

B= 2.5679e-004 *slope*C= -4.9624e-002 *intercept*

Rho= 0.9998686

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/d</i>
0.00	0.000	0.027	0.027	297	0.000	297					=
0.05	0.050					205	0				3
0.20	0.200	0.206	0.006	997	0.0 %	997					-4
2.50	2.500	2.408	-0.092	9571	0.0 %	9571					2
5.00	5.000	5.100	0.100	20054	0.0 %	20054					-1
7.50	7.500	7.453	-0.047	29216	0.0 %	29216					0
10.0	10.000	10.006	0.006	39160	0.0 %	39160					

LB117548
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	297	-		0 SFAM01.1	PPB	11/26/2021 13:48	Std
	0.2 S01	997	-		0.2 SFAM01.1	PPB	11/26/2021 13:51	Std
	2.5 S02	9571	-		2.5 SFAM01.1	PPB	11/26/2021 13:53	Std
	5 S03	20054	-		5 SFAM01.1	PPB	11/26/2021 13:55	Std
	7.5 S04	29216	-		7.5 SFAM01.1	PPB	11/26/2021 13:57	Std
	10 S05	39160	-		10 SFAM01.1	PPB	11/26/2021 14:00	Std
ICV091	ICV091	15815	4.0115	-	SFAM01.1	PPB	11/26/2021 14:02	SMPL
ICB091	ICB091	234	0.0105	-	SFAM01.1	PPB	11/26/2021 14:05	SMPL
CCV026	CCV026	19868	5.0522	-	SFAM01.1	PPB	11/26/2021 14:07	SMPL
CCB026	CCB026	223	0.0076	-	SFAM01.1	PPB	11/26/2021 14:09	SMPL
PB141032BL	PBW032	243	0.0128	-	SFAM01.1	PPB	11/26/2021 14:11	SMPL
M4759-01	SY-3-20211117	645	0.116	-	SFAM01.1	PPB	11/26/2021 14:14	SMPL
M4759-01S	SY-3-20211117S	4196	1.0279	-	SFAM01.1	PPB	11/26/2021 14:16	SMPL
M4759-01D	SY-3-20211117D	636	0.1137	-	SFAM01.1	PPB	11/26/2021 14:24	SMPL
M4759-02	SY-3-20211117	276	0.0212	-	SFAM01.1	PPB	11/26/2021 14:26	SMPL
M4775-02	PK-10S-20211118	230	0.0094	-	SFAM01.1	PPB	11/26/2021 14:29	SMPL
M4775-03	PK-10S-20211118	233	0.0102	-	SFAM01.1	PPB	11/26/2021 14:31	SMPL
M4775-04	PK-10I-20211118	486	0.0752	-	SFAM01.1	PPB	11/26/2021 14:33	SMPL
M4775-05	PK-10I-20211118	318	0.032	-	SFAM01.1	PPB	11/26/2021 14:35	SMPL
M4775-06	PK-10D-20211118	456	0.0675	-	SFAM01.1	PPB	11/26/2021 14:38	SMPL
M4775-07	PK-10D-20211118	372	0.0459	-	SFAM01.1	PPB	11/26/2021 14:40	SMPL
M4775-08	RW-12I-20211118	340	0.0377	-	SFAM01.1	PPB	11/26/2021 14:42	SMPL
M4775-09	RW-12I-20211118	297	0.0266	-	SFAM01.1	PPB	11/26/2021 14:44	SMPL
M4775-10	RW-12D-20211118	308	0.0295	-	SFAM01.1	PPB	11/26/2021 14:47	SMPL
M4775-11	RW-12D-20211118	305	0.0287	-	SFAM01.1	PPB	11/26/2021 14:49	SMPL
CCV027	CCV027	20781	5.2867	-	SFAM01.1	PPB	11/26/2021 14:51	SMPL
CCB027	CCB027	273	0.0205	-	SFAM01.1	PPB	11/26/2021 14:54	SMPL