


LB11755D

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2602e-004 slope

C= -9.4191e-002 y-intercept

Rho= 0.9999237

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.00	0.000	-0.013	-0.013	358	0.000	358					
0.05	0.050					205	0				6
0.20	0.200	0.212	0.012	1356	0.0 %	1356					1
2.50	2.500	2.521	0.021	11570	0.0 %	11570					0
5.00	5.000	5.020	0.020	22627	0.0 %	22627					-1
7.50	7.500	7.408	-0.092	33191	0.0 %	33191					1
10.0	10.000	10.053	0.053	44894	0.0 %	44894					

LB117550
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	358	-		0 SFAM01.1	PPB	11/26/2021 15:36	Std
	0.2 S01	1356	-		0.2 SFAM01.1	PPB	11/26/2021 15:40	Std
	2.5 S02	11570	-		2.5 SFAM01.1	PPB	11/26/2021 15:45	Std
	5 S03	22627	-		5 SFAM01.1	PPB	11/26/2021 15:47	Std
	7.5 S04	33191	-		7.5 SFAM01.1	PPB	11/26/2021 15:50	Std
	10 S05	44894	-		10 SFAM01.1	PPB	11/26/2021 15:52	Std
ICV092	ICV092	17809	3.931 -		SFAM01.1	PPB	11/26/2021 15:55	SMPL
ICB092	ICB092	311	-0.0239 -		SFAM01.1	PPB	11/26/2021 15:57	SMPL
CCV028	CCV028	22504	4.9921 -		SFAM01.1	PPB	11/26/2021 15:59	SMPL
CCB028	CCB028	305	-0.0253 -		SFAM01.1	PPB	11/26/2021 16:01	SMPL
PB141033BL	PBS033	433	0.0037 -		SFAM01.1	PPB	11/26/2021 16:04	SMPL
M4634-01	GB8T4	1159	0.1678 -		SFAM01.1	PPB	11/26/2021 16:06	SMPL
M4634-02	GB8T4D	1152	0.1662 -		SFAM01.1	PPB	11/26/2021 16:08	SMPL
M4634-03	GB8T4S	12737	2.7846 -		SFAM01.1	PPB	11/26/2021 16:10	SMPL
M4634-04	GB8T5	457	0.0091 -		SFAM01.1	PPB	11/26/2021 16:13	SMPL
M4634-05	GB8T6	918	0.1133 -		SFAM01.1	PPB	11/26/2021 16:15	SMPL
M4634-06	MGB8T7	22968	5.097 -		SFAM01.1	PPB	11/26/2021 16:17	SMPL
CCV029	CCV029	22648	5.0247 -		SFAM01.1	PPB	11/26/2021 16:26	SMPL
CCB029	CCB029	278	-0.0314 -		SFAM01.1	PPB	11/26/2021 16:28	SMPL