

Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB117556

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP94086
calibration std. phosphate 0.5 ppm	WP94085
calibration std. phosphate 0.3 ppm	WP94084
calibration std. phosphate 0.1 ppm	WP94083
calibration std. phosphate 0.05 ppm	WP94082
calibration std. 0 ppm	WP94081
phosphate CCV std.	WP94087
Combined reagent	WP94090
Sodium hydroxide, 1N	WP92662
Phosphate ICV-LCS Std	WP94088

Intercept: -0.0006

Slope: 0.6521

Regression: 0.99998

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.001		11/29/2021	11:40
2	CAL2	0.05	1	50	50	0.032	0.05	0	11/29/2021	11:40
3	CAL3	0.10	1	50	50	0.066	0.102	-2	11/29/2021	11:41
4	CAL4	0.30	1	50	50	0.192	0.295	1.7	11/29/2021	11:41
5	CAL5	0.50	1	50	50	0.326	0.501	-0.2	11/29/2021	11:42
6	CAL6	1.00	1	50	50	0.652	1.001	-0.1	11/29/2021	11:42



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB117556

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.493	11/29/2021	11:44
2	ICB		1	50	50	0.000	0.001	11/29/2021	11:44
3	CCV1	0.50	1	50	50	0.324	0.498	11/29/2021	11:45
4	CCB1		1	50	50	0.000	0.001	11/29/2021	11:45
5	PB141040BL		1	50	50	0.001	0.002	11/29/2021	11:46
6	PB141040BS	0.50	1	50	50	0.319	0.490	11/29/2021	11:46
7	M4785-01		1	50	50	1.502	2.304	11/29/2021	11:47
8	M4785-01DUP		1	50	50	1.503	2.306	11/29/2021	11:47
9	M4785-01MS	0.50	1	50	50	1.621	2.487	11/29/2021	11:48
10	M4785-01MSD	0.50	1	50	50	1.614	2.476	11/29/2021	11:48
11	M4834-01		1	50	50	1.025	1.573	11/29/2021	11:49
12	M4835-01		1	50	50	0.572	0.878	11/29/2021	11:50
13	M4785-01		10	50	50	0.584	0.896	11/29/2021	11:50
14	M4785-01DUP		10	50	50	0.586	0.900	11/29/2021	11:51
15	CCV2	0.50	1	50	50	0.327	0.502	11/29/2021	11:52
16	CCB2		1	50	50	0.001	0.002	11/29/2021	11:52
17	M4785-01MS	0.50	10	50	50	0.618	0.949	11/29/2021	11:53
18	M4785-01MSD	0.50	10	50	50	0.617	0.947	11/29/2021	11:53
19	M4834-01		10	50	50	0.331	0.509	11/29/2021	11:54
20	CCV3	0.50	1	50	50	0.325	0.499	11/29/2021	11:55
21	CCB3		1	50	50	0.000	0.001	11/29/2021	11:55