



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117561

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	11/26/2021	09:58
2	M4807-04		1	10	100.00	>212.0	>212.0	11/26/2021	10:20
3	M4807-04DUP		1	10	100.00	>212.0	>212.0	11/26/2021	10:45
4	M4817-01		1	15	100.00	>212.0	>212.0	11/26/2021	11:10
5	M4830-10		1	12	100.00	>212.0	>212.0	11/26/2021	11:40
6	M4837-16		1	13	100.00	>212.0	>212.0	11/26/2021	12:05
7	M4837-17		1	14	100.00	>212.0	>212.0	11/26/2021	12:35
8	M4837-18		1	10	100.00	>212.0	>212.0	11/26/2021	13:05
9	M4837-19		1	12	100.00	>212.0	>212.0	11/26/2021	13:40
10	M4837-20		1	14	100.00	>212.0	>212.0	11/26/2021	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117561

WorkList Name :		fp112421		WorkList ID :		154645		Department :		Wet-Chemistry		Date :		11-24-2021 09:01:55	
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method						
11/29/2021	Water	M4807-04	Flash Point	Cool 4 deg C	PSEG03	F11	TOT-COMPOSITE	11/22/2021	1010B						
11/30/2021	Water	M4817-01	Flash Point	Cool 4 deg C	PSEG03	F12	A6879	11/23/2021	1010B						
12/01/2021	Water	M4830-10	Flash Point	Cool 4 deg C	PSEG03	F21	30930	11/24/2021	1010B						
11/19/2021	Water	M4837-16	Flash Point	Cool 4 deg C	PSEG03	F21	50-26-29	11/24/2021	1010B						
11/19/2021	Water	M4837-17	Flash Point	Cool 4 deg C	PSEG03	F21	30-35	11/24/2021	1010B						
11/19/2021	Water	M4837-18	Flash Point	Cool 4 deg C	PSEG03	F21	20873	11/24/2021	1010B						
11/17/2021	Water	M4837-19	Flash Point	Cool 4 deg C	PSEG03	F21	20936	11/24/2021	1010B						
11/17/2021	Water	M4837-20	Flash Point	Cool 4 deg C	PSEG03	F21	20937	11/24/2021	1010B						

Date/Time 11/26/21 09:30
 Raw Sample Received by: PS(wc)
 Raw Sample Relinquished by: PS(wc)

Date/Time 11/26/21 14:15
 Raw Sample Received by: PS(wc)
 Raw Sample Relinquished by: PS(wc)