

LB 117562

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : 12 Instrument ID : Konelab

11/29/2021 14:16

Test: Phenolics-

Sample Id	Result <i>mg/L</i>	Dil. 1 +	Response	Errors
ICV1	0.990	0.0	0.106	
ICB1	0.003	0.0	0.004	
CCV1	0.984	0.0	0.106	
CCB1	0.004	0.0	0.004	
PB140929BL	0.005	0.0	0.004	
PB140929BS	1.004	0.0	0.108	
M4759-01	0.899	0.0	0.097	
M4759-01DUP	0.903	0.0	0.098	
M4759-01MS	1.852	0.0	0.196	
M4759-01MSD	1.856	0.0	0.196	
M4775-02	0.002	0.0	0.004	
M4775-04	0.200	0.0	0.025	
M4775-06	0.001	0.0	0.004	
M4775-08	0.040	0.0	0.008	
M4775-10	0.002	0.0	0.004	
DOC-1	0.974	0.0	0.105	
DOC-2	0.985	0.0	0.106	
DOC-3	0.973	0.0	0.105	
DOC-4	0.986	0.0	0.106	
CCV2	0.979	0.0	0.105	
CCB2	0.004	0.0	0.004	

N	21
Mean	0.650
SD	0.6092
CV%	93.76

Aquakem v. 7.2AQ1

Results from time period:

Mon Nov 29 12:56:54 2021

Mon Nov 29 13:52:27 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0046	mg/l	11/29/2021 12:56:54	
0.05PPM	A	Phenolics-	P	0.0504	mg/l	11/29/2021 12:56:55	
0.1PPM	A	Phenolics-	P	0.102	mg/l	11/29/2021 12:56:56	
0.25PPM	A	Phenolics-	P	0.2513	mg/l	11/29/2021 12:56:57	
0.50PPM	A	Phenolics-	P	0.5015	mg/l	11/29/2021 12:56:58	
1.0PPM	A	Phenolics-	P	0.9818	mg/l	11/29/2021 12:56:59	
2.0PPM	A	Phenolics-	P	2.0085	mg/l	11/29/2021 12:57:00	
ICV1	S	Phenolics-	P	0.9896	mg/l	11/29/2021 13:39:41	
ICB1	S	Phenolics-	P	0.0034	mg/l	11/29/2021 13:39:43	
CCV1	S	Phenolics-	P	0.9842	mg/l	11/29/2021 13:39:44	
CCB1	S	Phenolics-	P	0.0036	mg/l	11/29/2021 13:39:46	
PB140929BL	S	Phenolics-	P	0.0047	mg/l	11/29/2021 13:39:48	
PB140929BS	S	Phenolics-	P	1.0044	mg/l	11/29/2021 13:39:49	
M4759-01	S	Phenolics-	P	0.8986	mg/l	11/29/2021 13:39:50	
M4759-01DUP	S	Phenolics-	P	0.9032	mg/l	11/29/2021 13:39:51	
M4759-01MS	S	Phenolics-	P	1.8516	mg/l	11/29/2021 13:49:39	
M4759-01MSD	S	Phenolics-	P	1.8555	mg/l	11/29/2021 13:49:40	
M4775-02	S	Phenolics-	P	0.0016	mg/l	11/29/2021 13:49:41	
M4775-04	S	Phenolics-	P	0.2002	mg/l	11/29/2021 13:49:42	
M4775-06	S	Phenolics-	P	0.0014	mg/l	11/29/2021 13:49:43	
M4775-08	S	Phenolics-	P	0.04	mg/l	11/29/2021 13:49:44	
M4775-10	S	Phenolics-	P	0.0023	mg/l	11/29/2021 13:49:45	
DOC-1	S	Phenolics-	P	0.9742	mg/l	11/29/2021 13:49:46	
DOC-2	S	Phenolics-	P	0.9854	mg/l	11/29/2021 13:49:47	
DOC-3	S	Phenolics-	P	0.9728	mg/l	11/29/2021 13:49:48	
DOC-4	S	Phenolics-	P	0.9857	mg/l	11/29/2021 13:49:49	
CCV2	S	Phenolics-	P	0.9791	mg/l	11/29/2021 13:49:50	
CCB2	S	Phenolics-	P	0.0037	mg/l	11/29/2021 13:52:27	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

11/29/2021 13:07

Test Phenolics-


Accepted 11/29/2021 13:07

Factor 9.655

Bias 0.004

Coeff. of det. 0.999863

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.004	0.0046	0.0000	0.8
2	Phenol-2	0.009	0.0504	0.0500	2.0
3	Phenol-2	0.015	0.1020	0.1000	0.5
4	Phenol-2	0.030	0.2513	0.2500	0.3
5	Phenol-2	0.056	0.5015	0.5000	-1.9
6	Phenol-2	0.106	0.9818	1.0000	0.2
7	Phenol-2	0.212	2.0085	2.0000	

12/11/2021