

LB 117570


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44287



A= 0.0000e+000

B= 2.2955e-004 *slope*C= -8.0254e-002 *y-intercept*

Rho= 0.9998750

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	514	0.000	514					=
0.05	0.050					205	0				-4
0.20	0.200	0.193	-0.007	1192	0.0 %	1192					1
2.50	2.500	2.525	0.025	11350	0.0 %	11350					-1
5.00	5.000	4.950	-0.050	21913	0.0 %	21913					-1
7.50	7.500	7.408	-0.092	32622	0.0 %	32622					-1
10.0	10.000	10.086	0.086	44287	0.0 %	44287					1

LB117570

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	514	-		0 SFAM01.1	PPB	11/30/2021 10:42	Std
	0.2 S01	1192	-		0.2 SFAM01.1	PPB	11/30/2021 10:44	Std
	2.5 S02	11350	-		2.5 SFAM01.1	PPB	11/30/2021 10:47	Std
	5 S03	21913	-		5 SFAM01.1	PPB	11/30/2021 10:49	Std
	7.5 S04	32622	-		7.5 SFAM01.1	PPB	11/30/2021 10:51	Std
	10 S05	44287	-		10 SFAM01.1	PPB	11/30/2021 11:01	Std
ICV093	ICV093	18066	4.0668 -		SFAM01.1	PPB	11/30/2021 11:04	SMPL
ICB093	ICB093	511	0.037 -		SFAM01.1	PPB	11/30/2021 11:07	SMPL
CCV030	CCV030	22084	4.9891 -		SFAM01.1	PPB	11/30/2021 11:09	SMPL
CCB030	CCB030	518	0.0387 -		SFAM01.1	PPB	11/30/2021 11:11	SMPL
PB141073BL	PBS073	332	-0.004 -		SFAM01.1	PPB	11/30/2021 11:16	SMPL
M4715-01	GB8W6	906	0.1277 -		SFAM01.1	PPB	11/30/2021 11:18	SMPL
M4715-02	GB8W6D	838	0.1121 -		SFAM01.1	PPB	11/30/2021 11:20	SMPL
M4715-03	GB8W6S	13084	2.9232 -		SFAM01.1	PPB	11/30/2021 11:23	SMPL
M4715-04	GB8W7	751	0.0921 -		SFAM01.1	PPB	11/30/2021 11:25	SMPL
M4715-05	GB8W8	632	0.0648 -		SFAM01.1	PPB	11/30/2021 11:27	SMPL
M4715-06	GB8W9	877	0.1211 -		SFAM01.1	PPB	11/30/2021 11:30	SMPL
M4715-07	GB8X0	300	-0.0114 -		SFAM01.1	PPB	11/30/2021 11:32	SMPL
M4715-08	GB8X1	360	0.0024 -		SFAM01.1	PPB	11/30/2021 11:34	SMPL
M4715-09	GB8X2	453	0.0237 -		SFAM01.1	PPB	11/30/2021 11:36	SMPL
M4715-10	GB8X3	507	0.0361 -		SFAM01.1	PPB	11/30/2021 11:39	SMPL
M4715-11	GB8X4	659	0.071 -		SFAM01.1	PPB	11/30/2021 11:41	SMPL
M4715-12	GB8X5	373	0.0054 -		SFAM01.1	PPB	11/30/2021 11:43	SMPL
M4715-13	GB8X6	416	0.0152 -		SFAM01.1	PPB	11/30/2021 11:45	SMPL
M4715-14	GB8X7	644	0.0676 -		SFAM01.1	PPB	11/30/2021 11:48	SMPL
M4715-15	GB8X8	586	0.0543 -		SFAM01.1	PPB	11/30/2021 11:50	SMPL
M4715-16	GB8X9	781	0.099 -		SFAM01.1	PPB	11/30/2021 11:52	SMPL
M4715-17	GB8Y0	503	0.0352 -		SFAM01.1	PPB	11/30/2021 11:54	SMPL
M4715-18	GB8Y1	522	0.0396 -		SFAM01.1	PPB	11/30/2021 11:57	SMPL
M4715-19	GB8Y2	521	0.0393 -		SFAM01.1	PPB	11/30/2021 11:59	SMPL
CCV031	CCV031	21843	4.9338 -		SFAM01.1	PPB	11/30/2021 12:01	SMPL
CCB031	CCB031	489	0.032 -		SFAM01.1	PPB	11/30/2021 12:04	SMPL
M4715-20	GB8Y3	951	0.138 -		SFAM01.1	PPB	11/30/2021 12:06	SMPL
M4715-21	GB8Y4	1013	0.1523 -		SFAM01.1	PPB	11/30/2021 12:08	SMPL
M4715-22	GB8Y5	878	0.1213 -		SFAM01.1	PPB	11/30/2021 12:10	SMPL
PB141074BL	PBS074	260	-0.0206 -		SFAM01.1	PPB	11/30/2021 12:13	SMPL
M4716-01	GB8Y6	1538	0.2728 -		SFAM01.1	PPB	11/30/2021 12:15	SMPL
M4716-02	GB8Y7	201	-0.0341 -		SFAM01.1	PPB	11/30/2021 12:17	SMPL
M4716-03	GB8Y7D	184	-0.038 -		SFAM01.1	PPB	11/30/2021 12:19	SMPL
M4716-04	GB8Y7S	11757	2.6186 -		SFAM01.1	PPB	11/30/2021 12:22	SMPL
M4716-05	GB8Y8	245	-0.024 -		SFAM01.1	PPB	11/30/2021 12:24	SMPL
M4716-06	GB8Y9	625	0.0632 -		SFAM01.1	PPB	11/30/2021 12:26	SMPL
M4716-07	GB8Z0	495	0.0334 -		SFAM01.1	PPB	11/30/2021 12:28	SMPL
M4716-08	GB8Z1	533	0.0421 -		SFAM01.1	PPB	11/30/2021 12:31	SMPL
M4716-09	GB8Z2	616	0.0611 -		SFAM01.1	PPB	11/30/2021 12:33	SMPL
M4716-10	GB8Z3	567	0.0499 -		SFAM01.1	PPB	11/30/2021 12:35	SMPL

LB117570
INSTRUMENT ID : CV1

M4716-11	GB8Z4	611	0.06 -	SFAM01.1	PPB	11/30/2021 12:38 SMPL
M4716-12	GB8Z5	427	0.0178 -	SFAM01.1	PPB	11/30/2021 12:40 SMPL
M4716-13	GB8Z6	593	0.0559 -	SFAM01.1	PPB	11/30/2021 12:42 SMPL
M4716-14	GB8Z7	541	0.0439 -	SFAM01.1	PPB	11/30/2021 12:44 SMPL
M4716-15	GB8Z8	632	0.0648 -	SFAM01.1	PPB	11/30/2021 12:47 SMPL
M4716-16	GB8Z9	720	0.085 -	SFAM01.1	PPB	11/30/2021 12:49 SMPL
CCV032	CCV032	22125	4.9985 -	SFAM01.1	PPB	11/30/2021 12:51 SMPL
CCB032	CCB032	416	0.0152 -	SFAM01.1	PPB	11/30/2021 12:53 SMPL
M4716-17	GB900	648	0.0685 -	SFAM01.1	PPB	11/30/2021 12:56 SMPL
M4716-18	GB901	710	0.0827 -	SFAM01.1	PPB	11/30/2021 12:58 SMPL
M4716-19	GB902	992	0.1475 -	SFAM01.1	PPB	11/30/2021 13:00 SMPL
M4716-20	GB903	1219	0.1996 -	SFAM01.1	PPB	11/30/2021 13:02 SMPL
M4716-21	GB904	1126	0.1782 -	SFAM01.1	PPB	11/30/2021 13:05 SMPL
M4716-22	GB905	150	-0.0458 -	SFAM01.1	PPB	11/30/2021 13:07 SMPL
PB141075BL	PBS075	385	0.0081 -	SFAM01.1	PPB	11/30/2021 13:09 SMPL
M4719-01	GB946	82	-0.0614 -	SFAM01.1	PPB	11/30/2021 13:12 SMPL
M4719-02	GB947	62	-0.066 -	SFAM01.1	PPB	11/30/2021 13:14 SMPL
M4719-03	GB948	1072	0.1658 -	SFAM01.1	PPB	11/30/2021 13:16 SMPL
M4719-04	GB948D	1007	0.1509 -	SFAM01.1	PPB	11/30/2021 13:21 SMPL
M4719-05	GB948S	12044	2.6844 -	SFAM01.1	PPB	11/30/2021 13:26 SMPL
M4719-06	GB949	602	0.0579 -	SFAM01.1	PPB	11/30/2021 13:28 SMPL
M4719-07	GB950	512	0.0373 -	SFAM01.1	PPB	11/30/2021 13:31 SMPL
M4719-08	GB951	317	-0.0075 -	SFAM01.1	PPB	11/30/2021 13:33 SMPL
M4719-09	GB952	127	-0.0511 -	SFAM01.1	PPB	11/30/2021 13:35 SMPL
M4719-10	GB953	467	0.0269 -	SFAM01.1	PPB	11/30/2021 13:37 SMPL
M4719-11	GB954	673	0.0742 -	SFAM01.1	PPB	11/30/2021 13:40 SMPL
M4719-12	GB955	692	0.0786 -	SFAM01.1	PPB	11/30/2021 13:42 SMPL
M4719-13	GB956	947	0.1371 -	SFAM01.1	PPB	11/30/2021 13:44 SMPL
CCV033	CCV033	22330	5.0456 -	SFAM01.1	PPB	11/30/2021 13:46 SMPL
CCB033	CCB033	414	0.0148 -	SFAM01.1	PPB	11/30/2021 13:49 SMPL
M4719-14	GB957	751	0.0921 -	SFAM01.1	PPB	11/30/2021 13:51 SMPL
M4719-15	GB958	278	-0.0164 -	SFAM01.1	PPB	11/30/2021 13:53 SMPL
M4719-16	GB959	696	0.0795 -	SFAM01.1	PPB	11/30/2021 13:56 SMPL
M4719-17	GB960	689	0.0779 -	SFAM01.1	PPB	11/30/2021 13:58 SMPL
M4719-18	GB961	160	-0.0435 -	SFAM01.1	PPB	11/30/2021 14:00 SMPL
M4719-19	GB962	128	-0.0509 -	SFAM01.1	PPB	11/30/2021 14:02 SMPL
M4719-20	GB963	128	-0.0509 -	SFAM01.1	PPB	11/30/2021 14:05 SMPL
M4719-21	GB964	380	0.007 -	SFAM01.1	PPB	11/30/2021 14:07 SMPL
M4719-22	GB965	425	0.0173 -	SFAM01.1	PPB	11/30/2021 14:09 SMPL
PB141076BL	PBS076	307	-0.0098 -	SFAM01.1	PPB	11/30/2021 14:15 SMPL
M4720-01	GB966	290	-0.0137 -	SFAM01.1	PPB	11/30/2021 14:17 SMPL
M4720-02	GB967	292	-0.0132 -	SFAM01.1	PPB	11/30/2021 14:19 SMPL
M4720-03	GB967D	155	-0.0447 -	SFAM01.1	PPB	11/30/2021 14:25 SMPL
M4720-04	GB967S	12499	2.7889 -	SFAM01.1	PPB	11/30/2021 14:27 SMPL
M4720-05	GB968	246	-0.0238 -	SFAM01.1	PPB	11/30/2021 14:29 SMPL
M4720-06	GB969	266	-0.0192 -	SFAM01.1	PPB	11/30/2021 14:32 SMPL
M4720-07	GB970	708	0.0823 -	SFAM01.1	PPB	11/30/2021 14:34 SMPL

LB117570
INSTRUMENT ID : CV1

M4720-08	GB971	363	0.0031 -	SFAM01.1	PPB	11/30/2021 14:36 SMPL
M4720-09	GB972	567	0.0499 -	SFAM01.1	PPB	11/30/2021 14:38 SMPL
M4720-10	GB973	755	0.0931 -	SFAM01.1	PPB	11/30/2021 14:41 SMPL
CCV034	CCV034	22320	5.0433 -	SFAM01.1	PPB	11/30/2021 14:43 SMPL
CCB034	CCB034	417	0.0155 -	SFAM01.1	PPB	11/30/2021 14:45 SMPL
M4720-11	GB974	394	0.0102 -	SFAM01.1	PPB	11/30/2021 14:47 SMPL
M4720-12	GB975	609	0.0595 -	SFAM01.1	PPB	11/30/2021 14:50 SMPL
M4720-13	GB976	247	-0.0236 -	SFAM01.1	PPB	11/30/2021 14:52 SMPL
M4720-14	GB977	453	0.0237 -	SFAM01.1	PPB	11/30/2021 14:54 SMPL
M4720-15	GB978	320	-0.0068 -	SFAM01.1	PPB	11/30/2021 14:57 SMPL
M4720-16	GB979	314	-0.0082 -	SFAM01.1	PPB	11/30/2021 14:59 SMPL
M4720-17	GB980	1030	0.1562 -	SFAM01.1	PPB	11/30/2021 15:01 SMPL
CCV035	CCV035	22563	5.0991 -	SFAM01.1	PPB	11/30/2021 15:03 SMPL
CCB035	CCB035	416	0.0152 -	SFAM01.1	PPB	11/30/2021 15:06 SMPL