

LB 117571

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

11/30/2021 9:24

Test: Total CN

Sample Id	Result <i>mp/L</i>	Dil. 1 +	Response	Errors
ICV1	95.876	0.0	0.075	
ICB1	1.680	0.0	0.001	
CCV1	243.905	0.0	0.191	
CCB1	0.561	0.0	0.001	
PB141014BL	0.844	0.0	0.001	
M4817-01	1.063	0.0	0.001	
M4817-01DUP	1.216	0.0	0.001	
CCV2	248.025	0.0	0.194	
CCB2	1.365	0.0	0.001	

N 9
Mean 66.060
SD 106.6137
CV% 161.39

Aquakem v. 7.2AQ1

Results from time period:

Tue Nov 30 08:51:55 2021

Tue Nov 30 09:19:43 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.952	µg/l	11/30/2021 8:51:55	
5.OPPBCN	A	Total CN	P	5.0835	µg/l	11/30/2021 8:51:56	
10PPBCN	A	Total CN	P	10.0436	µg/l	11/30/2021 8:51:57	
50PPBCN	A	Total CN	P	49.0047	µg/l	11/30/2021 8:51:58	
100PPBCN	A	Total CN	P	99.4733	µg/l	11/30/2021 8:51:59	
250PPBCN	A	Total CN	P	250.4794	µg/l	11/30/2021 8:52:00	
500PPBCN	A	Total CN	P	499.9635	µg/l	11/30/2021 8:52:01	
ICV1	S	Total CN	P	95.8757	µg/l	11/30/2021 9:16:19	
ICB1	S	Total CN	P	1.6802	µg/l	11/30/2021 9:16:20	
CCV1	S	Total CN	P	243.9048	µg/l	11/30/2021 9:16:21	
CCB1	S	Total CN	P	0.5609	µg/l	11/30/2021 9:16:22	
PB141014BL	S	Total CN	P	0.8441	µg/l	11/30/2021 9:16:23	
M4817-01	S	Total CN	P	1.0631	µg/l	11/30/2021 9:16:24	
M4817-01DUPS	S	Total CN	P	1.216	µg/l	11/30/2021 9:16:25	
CCV2	S	Total CN	P	248.0254	µg/l	11/30/2021 9:16:26	
CCB2	S	Total CN	P	1.3653	µg/l	11/30/2021 9:16:27	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

11/30/2021 8:52

Test Total CN


Accepted 11/30/2021 8:52

Factor 1276

Bias 0

Coeff. of det. 0.999988

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.9520	0.0000	1.7
2	5.0PPBCN	0.004	5.0835	5.0000	0.4
3	10PPBCN	0.008	10.0436	10.0000	-2.0
4	50PPBCN	0.038	49.0047	50.0000	-0.5
5	100PPBCN	0.078	99.4733	100.0000	0.2
6	250PPBCN	0.196	250.4794	250.0000	0.0
7	500PPBCN	0.392	499.9635	500.0000	

11/30/21
12