

CB 117579

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

11/30/2021 11:31

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.950	0.0	0.077	
ICB1	1.305	0.0	0.001	
CCV1	249.921	0.0	0.196	
CCB1	1.310	0.0	0.001	
PB141058BL	1.395	0.0	0.001	
PB141058BS	99.903	0.0	0.078	
M4806-07	10.217	0.0	0.008	
M4806-07DUP	9.770	0.0	0.008	
M4820-07	10.553	0.0	0.008	
CCV2	268.151	0.0	0.210	
CCB2	1.490	0.0	0.001	

N 11
Mean 68.360
SD 101.4173
CV% 148.36

Aquakem v. 7.2AQ1

Results from time period:

Tue Nov 30 11:25:00 2021

Tue Nov 30 11:29:22 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.952	µg/l	11/30/2021 8:51:55	
5.OPPBCN	A	Total CN	P	5.0835	µg/l	11/30/2021 8:51:56	
10PPBCN	A	Total CN	P	10.0436	µg/l	11/30/2021 8:51:57	
50PPBCN	A	Total CN	P	49.0047	µg/l	11/30/2021 8:51:58	
100PPBCN	A	Total CN	P	99.4733	µg/l	11/30/2021 8:51:59	
250PPBCN	A	Total CN	P	250.4794	µg/l	11/30/2021 8:52:00	
500PPBCN	A	Total CN	P	499.9635	µg/l	11/30/2021 8:52:01	
ICV1	S	Total CN	P	97.95	µg/l	11/30/2021 11:25:01	
ICB1	S	Total CN	P	1.3051	µg/l	11/30/2021 11:25:02	
CCV1	S	Total CN	P	249.9215	µg/l	11/30/2021 11:25:05	
CCB1	S	Total CN	P	1.3095	µg/l	11/30/2021 11:25:07	
PB141058BL	S	Total CN	P	1.395	µg/l	11/30/2021 11:25:08	
PB141058BS	S	Total CN	P	99.9025	µg/l	11/30/2021 11:25:09	
M4806-07	S	Total CN	P	10.2168	µg/l	11/30/2021 11:25:10	
M4806-07DUP	S	Total CN	P	9.7695	µg/l	11/30/2021 11:29:19	
M4820-07	S	Total CN	P	10.5529	µg/l	11/30/2021 11:29:20	
CCV2	S	Total CN	P	268.1508	µg/l	11/30/2021 11:29:21	
CCB2	S	Total CN	P	1.49	µg/l	11/30/2021 11:29:22	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

11/30/2021 8:52


Test Total CN

Accepted 11/30/2021 8:52

Factor 1276
Bias 0

Coeff. of det. 0.999988

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.9520	0.0000	1.7
2	5.0PPBCN	0.004	5.0835	5.0000	0.4
3	10PPBCN	0.008	10.0436	10.0000	-2.0
4	50PPBCN	0.038	49.0047	50.0000	-0.5
5	100PPBCN	0.078	99.4733	100.0000	0.2
6	250PPBCN	0.196	250.4794	250.0000	0.0
7	500PPBCN	0.392	499.9635	500.0000	

11/30/21
12