



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117582

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/29/2021	11:15
2	CAL2	1	Water	NA	NA	20.3	7.00	11/29/2021	11:16
3	CAL3	1	Water	NA	NA	20.3	10.03	11/29/2021	11:19
4	ICV	1	Water	NA	NA	20.5	7.02	11/29/2021	11:20
5	CCV1	1	Water	NA	NA	20.4	2.01	11/29/2021	11:22
6	M4721-01	1	Solid	20.03	20	21.7	9.10	11/29/2021	11:25
7	M4721-01DUP	1	Solid	20.04	20	21.8	9.11	11/30/2021	11:26
8	M4721-02	1	Solid	20.02	20	21.9	8.51	11/29/2021	11:30
9	M4721-03	1	Solid	20.03	20	21.2	8.71	11/29/2021	11:33
10	CCV2	1	Water	NA	NA	20.4	12.03	11/29/2021	11:35

VB 117582

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S M4721 WorkList ID : 154698 Department : Wet-Chemistry Date : 11-29-2021 11:09:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/23/2021	Solid	M4721-01	F pH	Cool 4 deg C	KAKS01	D32	SAMPLE-1	11/23/2021	9045D
11/23/2021	Solid	M4721-02	F pH	Cool 4 deg C	KAKS01	D32	SAMPLE-2	11/23/2021	9045D
11/23/2021	Solid	M4721-03	F pH	Cool 4 deg C	KAKS01	D32	SAMPLE-3	11/23/2021	9045D

Date/Time 11.29.21 11:10
Raw Sample Received by: JP (WV)
Raw Sample Relinquished by: JP SM

Date/Time 11.29.21 13:00
Raw Sample Received by: JP SM
Raw Sample Relinquished by: JP (WV)