



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Sohil

Run Number: LB117589

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/30/2021	16:10
2	CAL2	1	Water	NA	NA	20.4	7.00	11/30/2021	16:11
3	CAL3	1	Water	NA	NA	20.4	10.03	11/30/2021	16:12
4	ICV	1	Water	NA	NA	20.6	6.98	11/30/2021	16:15
5	CCV1	1	Water	NA	NA	20.5	2.01	11/30/2021	16:20
6	M4863-08	1	Solid	20.03	20	22.4	10.11	11/30/2021	16:25
7	M4863-08DUP	1	Solid	20.04	20	22.5	10.12	11/30/2021	16:26
8	CCV2	1	Water	NA	NA	20.5	12.03	11/30/2021	16:30

WORKLIST(Hardcopy Internal Chain)

18117589

WorkList Name : CORROSIONITY M4863 WorkList ID : 154744 Department : Wet-Chemistry Date : 11-30-2021 15:46:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/30/2021	Solid	M4863-08	Corrosivity	Cool 4 deg C	DVIR01	F31	C-6	11/30/2021	9045D

Date/Time 11/30/21 15:50
 Raw Sample Received by: JR (SM)
 Raw Sample Relinquished by: JR (SM)

Date/Time 11/30/21 18:00
 Raw Sample Received by: JR (SM)
 Raw Sample Relinquished by: JR (SM)