

LB117590


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

39694



A= 0.0000e+000

B= 2.5039e-004

C= 2.2618e-002

Rho= 0.9998973

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.034	0.034	46	0.000	46					-
0.2	0.200	0.196	-0.004	691	0.0 %	691					-2
2.5	2.500	2.469	-0.031	9769	0.0 %	9769					-1
5.0	5.000	4.941	-0.059	19644	0.0 %	19644					-1
7.5	7.500	7.598	0.098	30256	0.0 %	30256					1
10.0	10.000	9.962	-0.038	39694	0.0 %	39694					0

LB117590 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 01 Dec 2021 10:26:38					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7470A	PPB	01 Dec 2021 10:31:35	S	Std
0.2 - 1	50.2	691	-	0.2000	7470A	PPB	01 Dec 2021 10:33:51	S	Std
2.5 - 1	52.5	9769	-	2.5000	7470A	PPB	01 Dec 2021 10:36:08	S	Std
5.0 - 1	55	19644	-	5.0000	7470A	PPB	01 Dec 2021 10:38:26	S	Std
7.5 - 1	57.5	30256	-	7.5000	7470A	PPB	01 Dec 2021 10:40:42	S	Std
10.0 - 1	510	39694	-	10.0000	7470A	PPB	01 Dec 2021 10:42:58	S	Std
ICV93 - 1	ICV93	15938	4.0134	-	7470A	PPB	01 Dec 2021 10:48:13	U	SMPL
ICB93 - 1	ICB93	-41	0.0124	-	7470A	PPB	01 Dec 2021 10:50:28	U	SMPL
CCV36 - 1	CCV36	19491	4.9030	-	7470A	PPB	01 Dec 2021 10:52:43	U	SMPL
CCB36 - 1	CCB36	-40	0.0126	-	7470A	PPB	01 Dec 2021 10:54:58	U	SMPL
CRA - 1	CRA	672	0.1909	-	7470A	PPB	01 Dec 2021 10:57:13	U	SMPL
HighStd - 1	HighStd	38851	9.7506	-	7470A	PPB	01 Dec 2021 10:59:28	U	SMPL
ChkStd - 1	ChkStd	28087	7.0554	-	7470A	PPB	01 Dec 2021 11:01:44	U	SMPL
PB141080BL - 1	PBW	-91	-0.0002	-	7470A	PPB	01 Dec 2021 11:04:01	U	SMPL
PB141080BS - 1	LCSW	14266	3.5947	-	7470A	PPB	01 Dec 2021 11:06:18	U	SMPL
M4848-02 - 1	PSC124055	-20	0.0176	-	7470A	PPB	01 Dec 2021 11:08:35	U	SMPL
M4848-04 - 1	PSC124019	30	0.0301	-	7470A	PPB	01 Dec 2021 11:10:51	U	SMPL
M4848-04DUP - 1	PSC124019DUP	38	0.0321	-	7470A	PPB	01 Dec 2021 11:13:08	U	SMPL
M4848-04MS - 1	PSC124019MS	15046	3.7900	-	7470A	PPB	01 Dec 2021 11:15:25	U	SMPL
M4848-04MSD - 1	PSC124019MSD	13893	3.5013	-	7470A	PPB	01 Dec 2021 11:17:42	U	SMPL
CCV37 - 1	CCV37	19318	4.8597	-	7470A	PPB	01 Dec 2021 11:19:57	U	SMPL
CCB37 - 1	CCB37	-25	0.0164	-	7470A	PPB	01 Dec 2021 11:22:12	U	SMPL
PB141081BL - 1	PBW	59	0.0374	-	7470A	PPB	01 Dec 2021 11:24:27	U	SMPL
PB141081BS - 1	LCSW	15361	3.8689	-	7470A	PPB	01 Dec 2021 11:29:16	U	SMPL
M4817-01 - 1	A6879	316	0.1017	-	7470A	PPB	01 Dec 2021 11:36:20	U	SMPL
M4817-01DUP - 1	A6879DUP	292	0.0957	-	7470A	PPB	01 Dec 2021 11:38:35	U	SMPL
M4817-01MS - 1	A6879MS	16743	4.2149	-	7470A	PPB	01 Dec 2021 11:40:51	U	SMPL
M4817-01MSD - 1	A6879MSD	16397	4.1283	-	7470A	PPB	01 Dec 2021 11:43:07	U	SMPL
PB141080TB - 1	PB141053TB	-22	0.0171	-	7470A	PPB	01 Dec 2021 11:45:23	U	SMPL
CCV38 - 1	CCV38	19141	4.8154	-	7470A	PPB	01 Dec 2021 11:47:40	U	SMPL
CCB38 - 1	CCB38	-26	0.0161	-	7470A	PPB	01 Dec 2021 11:49:57	U	SMPL
M4848-04LX5 - 1		34	0.0311	-	7470A	PPB	01 Dec 2021 11:52:12	U	SMPL
M4848-04A - 1		16424	4.1351	-	7470A	PPB	01 Dec 2021 11:54:28	U	SMPL
M4817-01LX5 - 1		194	0.0712	-	7470A	PPB	01 Dec 2021 12:04:55	U	SMPL
M4817-01A - 1		16213	4.0822	-	7470A	PPB	01 Dec 2021 12:07:10	U	SMPL
CCV39 - 1	CCV39	19369	4.8725	-	7470A	PPB	01 Dec 2021 12:09:25	U	SMPL
CCB39 - 1	CCB39	-51	0.0098	-	7470A	PPB	01 Dec 2021 12:11:41	U	SMPL