



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117594

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	12/01/2021	12:40
2	M4866-03		1	12	100.00	>212.0	>212.0	12/01/2021	12:58
3	M4866-03DUP		1	12	100.00	>212.0	>212.0	12/01/2021	13:20
4	M4866-04		1	15	100.00	>212.0	>212.0	12/01/2021	13:45
5	M4866-05		1	14	100.00	>212.0	>212.0	12/01/2021	14:10
6	M4866-06		1	13	100.00	>212.0	>212.0	12/01/2021	14:35
7	M4866-07		1	15	100.00	>212.0	>212.0	12/01/2021	14:55
8	M4866-08		1	12	100.00	>212.0	>212.0	12/01/2021	15:22
9	M4866-09		1	14	100.00	>212.0	>212.0	12/01/2021	15:42
10	M4866-10		1	13	100.00	>212.0	>212.0	12/01/2021	16:05
11	M4866-11		1	12	100.00	>212.0	>212.0	12/01/2021	16:32
12	M4866-12		1	15	100.00	>212.0	>212.0	12/01/2021	16:58
13	M4866-17		1	14	100.00	>212.0	>212.0	12/01/2021	17:25
14	M4866-21		1	14	100.00	>212.0	>212.0	12/01/2021	17:45
15	M4866-22		1	15	100.00	>212.0	>212.0	12/01/2021	18:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp 12121

WorkList ID : 154770

Department : Wet-Chemistry

Date : 12-01-2021 10:46:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/06/2021	Water	M4866-03	Flash Point	Cool 4 deg C	PSEG03	F32	30913	11/30/2021	1010B
12/06/2021	Water	M4866-04	Flash Point	Cool 4 deg C	PSEG03	F32	40468	11/30/2021	1010B
12/06/2021	Water	M4866-05	Flash Point	Cool 4 deg C	PSEG03	F32	40467	11/30/2021	1010B
12/06/2021	Water	M4866-06	Flash Point	Cool 4 deg C	PSEG03	F32	16941	11/30/2021	1010B
12/06/2021	Water	M4866-07	Flash Point	Cool 4 deg C	PSEG03	F32	16985	11/30/2021	1010B
12/06/2021	Water	M4866-08	Flash Point	Cool 4 deg C	PSEG03	F32	16984	11/30/2021	1010B
12/06/2021	Water	M4866-09	Flash Point	Cool 4 deg C	PSEG03	F32	30938	11/30/2021	1010B
12/06/2021	Water	M4866-10	Flash Point	Cool 4 deg C	PSEG03	F32	30914	11/30/2021	1010B
12/06/2021	Water	M4866-11	Flash Point	Cool 4 deg C	PSEG03	F32	30911	11/30/2021	1010B
12/06/2021	Water	M4866-12	Flash Point	Cool 4 deg C	PSEG03	F32	30912	11/30/2021	1010B
12/06/2021	Water	M4866-17	Flash Point	Cool 4 deg C	PSEG03	F32	DRUM-COMP	11/30/2021	1010B
12/06/2021	Water	M4866-21	Flash Point	Cool 4 deg C	PSEG03	F32	TOTE-COMP	11/30/2021	1010B
12/06/2021	Water	M4866-22	Flash Point	Cool 4 deg C	PSEG03	F32	16995	11/30/2021	1010B

Date/Time 12/01/21 12:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: SP (WC)

Date/Time 12/01/21 18:15
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: KS (WC)