

LB 117599


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55369



A= 0.0000e+000

B= 1.8160e-004

C= 2.1460e-002

Rho= 0.9998029

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.052	0.052	170	0.000	170					
0.2	0.200	0.226	0.026	1127	0.0 %	1127					13
2.5	2.500	2.404	-0.096	13121	0.0 %	13121					-4
5.0	5.000	5.048	0.048	27677	0.0 %	27677					1
7.5	7.500	7.393	-0.107	40591	0.0 %	40591					-1
10.0	10.000	10.077	0.077	55369	0.0 %	55369					1

LB117599 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Dec 2021 14:41:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	170	-	0.0000	7471B	PPB	01 Dec 2021 14:43:33	S	Std
0.2 - 1	50.2	1127	-	0.2000	7471B	PPB	01 Dec 2021 14:45:50	S	Std
2.5 - 1	52.5	13121	-	2.5000	7471B	PPB	01 Dec 2021 14:48:07	S	Std
5.0 - 1	55	27677	-	5.0000	7471B	PPB	01 Dec 2021 14:50:24	S	Std
7.5 - 1	57.5	40591	-	7.5000	7471B	PPB	01 Dec 2021 14:52:40	S	Std
10.0 - 1	510	55369	-	10.0000	7471B	PPB	01 Dec 2021 14:54:57	S	Std
ICV94 - 1	ICV94	21141	4.0533	-	7471B	PPB	01 Dec 2021 14:57:40	U	SMPL
ICB94 - 1	ICB94	87	-0.0441	-	7471B	PPB	01 Dec 2021 14:59:55	U	SMPL
CCV40 - 1	CCV40	26810	5.1566	-	7471B	PPB	01 Dec 2021 15:02:10	U	SMPL
CCB40 - 1	CCB40	95	-0.0425	-	7471B	PPB	01 Dec 2021 15:04:25	U	SMPL
CRA - 1	CRA	1108	0.2227	-	7471B	PPB	01 Dec 2021 15:09:21	U	SMPL
HighStd - 1	HighStd	53509	9.7389	-	7471B	PPB	01 Dec 2021 15:11:36	U	SMPL
ChkStd - 1	ChkStd	36820	6.7081	-	7471B	PPB	01 Dec 2021 15:13:51	U	SMPL
PB141098BL - 1	PBS	93	0.0383	-	7471B	PPB	01 Dec 2021 15:16:08	U	SMPL
PB141098BS - 1	LCSS	22070	4.0294	-	7471B	PPB	01 Dec 2021 15:18:24	U	SMPL
M4721-01 - 1	SAMPLE-1	1073	0.2163	-	7471B	PPB	01 Dec 2021 15:20:41	U	SMPL
M4721-01DUP - 1	SAMPLE-1DUP	1083	0.2181	-	7471B	PPB	01 Dec 2021 15:25:31	U	SMPL
M4721-01MS - 1	SAMPLE-1MS	23523	4.2933	-	7471B	PPB	01 Dec 2021 15:27:50	U	SMPL
M4721-01MSD - 1	SAMPLE-1MSD	23123	4.2207	-	7471B	PPB	01 Dec 2021 15:30:07	U	SMPL
M4721-02 - 1	SAMPLE-2	59927	10.9044	-	7471B	PPB	01 Dec 2021 15:32:23	U	SMPL
CCV41 - 1	CCV41	27032	4.9306	-	7471B	PPB	01 Dec 2021 15:34:38	U	SMPL
CCB41 - 1	CCB41	89	0.0376	-	7471B	PPB	01 Dec 2021 15:36:55	U	SMPL
M4721-03 - 1	SAMPLE-3	11190	2.0536	-	7471B	PPB	01 Dec 2021 15:39:10	U	SMPL
M4848-01 - 1	PSC124055	1559	0.3046	-	7471B	PPB	01 Dec 2021 15:41:25	U	SMPL
M4848-03 - 1	PSC124019	1648	0.3207	-	7471B	PPB	01 Dec 2021 15:43:41	U	SMPL
M4849-01 - 1	HD-1-112921	1834	0.3545	-	7471B	PPB	01 Dec 2021 15:45:57	U	SMPL
M4855-01 - 1	LS-RO-COMP	1082	0.2180	-	7471B	PPB	01 Dec 2021 15:48:13	U	SMPL
M4860-01 - 1	BBC	5606	1.0395	-	7471B	PPB	01 Dec 2021 15:50:29	U	SMPL
M4865-01 - 1	TR-05-113021	1463	0.2871	-	7471B	PPB	01 Dec 2021 15:52:45	U	SMPL
M4866-01 - 1	41679	344	0.0839	-	7471B	PPB	01 Dec 2021 15:55:02	U	SMPL
M4866-02 - 1	TOTE-SLUDGE	47540	8.6549	-	7471B	PPB	01 Dec 2021 15:57:19	U	SMPL
CCV42 - 1	CCV42	27586	5.0312	-	7471B	PPB	01 Dec 2021 15:59:35	U	SMPL
CCB42 - 1	CCB42	101	0.0398	-	7471B	PPB	01 Dec 2021 16:01:52	U	SMPL
M4721-01LX5 - 1		323	0.0801	-	7471B	PPB	01 Dec 2021 16:04:07	U	SMPL
M4721-01A - 1		23793	4.3423	-	7471B	PPB	01 Dec 2021 16:06:23	U	SMPL
M4721-02DLX2 - 1		30407	5.5435	-	7471B	PPB	01 Dec 2021 16:08:38	U	SMPL
CCV43 - 1	CCV43	27423	5.0016	-	7471B	PPB	01 Dec 2021 16:10:54	U	SMPL
CCB43 - 1	CCB43	99	0.0394	-	7471B	PPB	01 Dec 2021 16:13:10	U	SMPL