



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 12/2/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 12/01/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 12/02/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117603

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4881-01	EW9D7	1	1.1	8.50	9.6	8.85	91.2	
M4881-02	EW9D8	2	1.1	8.51	9.61	7.63	76.7	
M4881-03	EW9D9	3	1.1	8.81	9.91	7.44	72.0	
M4881-04	EW9E0	4	1.09	8.89	9.98	8.41	82.3	
M4881-05	EW9E1	5	1.09	8.55	9.64	8.64	88.3	
M4881-06	EW9E2	6	1.09	8.66	9.75	7.96	79.3	
M4881-07	EW9E3	7	1.1	8.85	9.95	8.6	84.7	
M4881-08	EW9E4	8	1.08	8.68	9.76	8.09	80.8	
M4881-09	EW9E5	9	1.09	8.65	9.74	8.85	89.7	
M4881-10	EW9E6	10	1.1	8.85	9.95	9.00	89.3	
M4881-11	EW9E7	11	1.09	8.87	9.96	8.93	88.4	
M4881-12	EW9E8	12	1.1	8.42	9.52	9.06	94.5	
M4881-13	EW9E9	13	1.09	8.71	9.8	6.88	66.5	
M4881-14	EW9F0	14	1.09	8.71	9.8	8.62	86.5	
M4881-15	EW9F1	15	1.07	8.90	9.97	8.96	88.7	
M4881-16	EW9F2	16	1.1	8.69	9.79	9.35	94.9	
M4881-17	EW9F3	17	1.1	8.77	9.87	9.48	95.6	
M4881-18	EW9F4	18	1.09	8.72	9.81	8.96	90.3	
M4881-19	EW9F5	19	1.07	8.72	9.79	7.92	78.6	
M4881-20	EW9F6	20	1.08	8.49	9.57	8.43	86.6	
M4881-21	EW9F6MS	21	1.08	8.49	9.57	8.43	86.6	
M4881-22	EW9F6MSD	22	1.08	8.49	9.57	8.43	86.6	
M4881-23	VHBLK001	23	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$