



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/3/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:10
In Date: 12/02/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:48
Out Date: 12/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117609

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4885-01	EW9K8	1	1.12	8.72	9.84	6.95	66.9	
M4885-02	EW9K9	2	1.1	8.80	9.9	9.00	89.8	
M4885-03	EW9L0	3	1.14	8.70	9.84	9.4	94.9	
M4885-04	EW9L1	4	1.13	8.64	9.77	9.33	94.9	
M4885-05	EW9L2	5	1.12	8.40	9.52	8.68	90.0	
M4885-06	EW9L3	6	1.17	8.50	9.67	7.69	76.7	
M4885-07	EW9L4	7	1.14	8.60	9.74	8.67	87.6	
M4885-08	EW9L5	8	1.15	8.82	9.97	9.45	94.1	
M4885-09	EW9L6	9	1.11	8.74	9.85	8.07	79.6	
M4885-10	EW9L7	10	1.16	8.50	9.66	8.23	83.2	
M4885-11	EW9L8	11	1.16	8.65	9.81	7.98	78.8	
M4885-12	EW9L9	12	1.14	8.52	9.66	8.19	82.7	
M4885-13	EW9M0	13	1.13	8.40	9.53	8.74	90.6	
M4885-14	EW9M1	14	1.14	8.58	9.72	8.52	86.0	
M4885-15	EW9M2	15	1.17	8.67	9.84	7.78	76.2	
M4885-16	EW9M3	16	1.19	8.71	9.9	8.25	81.1	
M4885-17	EW9M4	17	1.16	8.81	9.97	8.35	81.6	
M4885-18	EW9M4MS	18	1.16	8.81	9.97	8.35	81.6	
M4885-19	EW9M4MSD	19	1.16	8.81	9.97	8.35	81.6	
M4885-20	EW9M5	20	1.17	8.51	9.68	8.51	86.3	
M4885-21	EW9M6	21	1.18	8.55	9.73	8.02	80.0	
M4885-22	EW9M7	22	1.1	8.57	9.67	7.98	80.3	
M4885-23	VHBLK001	23	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$