


LB117611

7470A

INSTRUMENT ID: CPT

 μ Abs.:

38721



Linear

A= 0.0000e+000

B= 2.5621e-004

C= 1.2325e-002

Rho= 0.9998108

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	43	0.000	43					—
0.2	0.200	0.206	0.006	754	0.0 %	754					3
2.5	2.500	2.384	-0.116	9256	0.0 %	9256					-5
5.0	5.000	5.092	0.092	19825	0.0 %	19825					2
7.5	7.500	7.563	0.063	29470	0.0 %	29470					1
10.0	10.000	9.933	-0.067	38721	0.0 %	38721					-1

LB117611 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Dec 2021 11:38:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	43	-	0.0000	7470A	PPB	02 Dec 2021 11:41:02	S	Std
0.2 - 1	SO2	754	-	0.2000	7470A	PPB	02 Dec 2021 11:43:20	S	Std
2.5 - 1	52.5	9256	-	2.5000	7470A	PPB	02 Dec 2021 11:45:37	S	Std
5.0 - 1	85	19825	-	5.0000	7470A	PPB	02 Dec 2021 11:47:54	S	Std
7.5 - 1	37.5	29470	-	7.5000	7470A	PPB	02 Dec 2021 11:50:10	S	Std
10.0 - 1	S10	38721	-	10.0000	7470A	PPB	02 Dec 2021 11:55:42	S	Std
ICV95 - 1	ICV95	14540	3.7376	-	7470A	PPB	02 Dec 2021 11:58:23	U	SMPL
ICB95 - 1	ICB95	-27	0.0054	-	7470A	PPB	02 Dec 2021 12:00:38	U	SMPL
CCV44 - 1	CCV44	19609	5.0363	-	7470A	PPB	02 Dec 2021 12:02:53	U	SMPL
CCB44 - 1	CCB44	-32	0.0041	-	7470A	PPB	02 Dec 2021 12:05:08	U	SMPL
CRA - 1	CRA	726	0.1983	-	7470A	PPB	02 Dec 2021 12:07:23	U	SMPL
HighStd - 1	HighStd	38843	9.9642	-	7470A	PPB	02 Dec 2021 12:09:38	U	SMPL
ChkStd - 1	ChkStd	26794	6.8772	-	7470A	PPB	02 Dec 2021 12:11:53	U	SMPL
PB141102BL - 1	PBW	-72	-0.0061	-	7470A	PPB	02 Dec 2021 12:14:10	U	SMPL
PB141102BS - 1	LCSW	14738	3.7883	-	7470A	PPB	02 Dec 2021 12:16:26	U	SMPL
M4855-03 - 1	LS-RO-COMP	-32	0.0041	-	7470A	PPB	02 Dec 2021 12:18:42	U	SMPL
M4863-08 - 1	C-6	24	0.0185	-	7470A	PPB	02 Dec 2021 12:20:58	U	SMPL
M4863-08DUP - 1	C-6DUP	32	0.0205	-	7470A	PPB	02 Dec 2021 12:23:15	U	SMPL
M4863-08MS - 1	C-6MS	15984	4.1076	-	7470A	PPB	02 Dec 2021 12:25:32	U	SMPL
M4863-08MSD - 1	C-6MSD	14591	3.7507	-	7470A	PPB	02 Dec 2021 12:27:48	U	SMPL
CCV45 - 1	CCV45	19979	5.1311	-	7470A	PPB	02 Dec 2021 12:30:03	U	SMPL
CCB45 - 1	CCB45	-42	0.0016	-	7470A	PPB	02 Dec 2021 12:32:18	U	SMPL
PB141103BL - 1	PBW	47	0.0244	-	7470A	PPB	02 Dec 2021 12:34:33	U	SMPL
PB141103BS - 1	LCSW	13059	3.3581	-	7470A	PPB	02 Dec 2021 12:36:48	U	SMPL
M4866-21 - 1	TOTE-COMP	28221	7.2428	-	7470A	PPB	02 Dec 2021 12:43:54	U	SMPL
M4866-21DUP - 1	TOTE-COMPDUP	27948	7.1728	-	7470A	PPB	02 Dec 2021 12:46:09	U	SMPL
M4866-21MS - 1	TOTE-COMPMS	44154	11.3249	-	7470A	PPB	02 Dec 2021 12:48:26	U	SMPL
M4866-21MSD - 1	TOTE-COMPMSD	41467	10.6365	-	7470A	PPB	02 Dec 2021 12:56:53	U	SMPL
PB141102TB - 1	PB141085TB	-87	-0.0100	-	7470A	PPB	02 Dec 2021 12:59:09	U	SMPL
CCV46 - 1	CCV46	19416	4.9869	-	7470A	PPB	02 Dec 2021 13:01:27	U	SMPL
CCB46 - 1	CCB46	-23	0.0064	-	7470A	PPB	02 Dec 2021 13:03:44	U	SMPL
M4863-08LX5 - 1		18	0.0169	-	7470A	PPB	02 Dec 2021 13:05:59	U	SMPL
M4863-08A - 1		14957	3.8444	-	7470A	PPB	02 Dec 2021 13:08:15	U	SMPL
M4866-21LX5 - 1		6502	1.6782	-	7470A	PPB	02 Dec 2021 13:10:30	U	SMPL
M4866-21A - 1		39114	10.0336	-	7470A	PPB	02 Dec 2021 13:15:23	U	SMPL
M4866-21DLX2 - 1		14644	3.7642	-	7470A	PPB	02 Dec 2021 13:17:38	U	SMPL
M4866-21DUPDLX2 - 1		14106	3.6264	-	7470A	PPB	02 Dec 2021 13:19:54	U	SMPL
M4866-21MSDLX2 - 1		22865	5.8705	-	7470A	PPB	02 Dec 2021 13:22:11	U	SMPL
M4866-21MSDDLX2 - 1		22426	5.7580	-	7470A	PPB	02 Dec 2021 13:24:26	U	SMPL
M4866-21LDLX10 - 1		2961	0.7710	-	7470A	PPB	02 Dec 2021 13:26:43	U	SMPL
M4866-21ADLX2 - 1		20136	5.1713	-	7470A	PPB	02 Dec 2021 13:28:59	U	SMPL
CCV47 - 1	CCV47	19421	4.9881	-	7470A	PPB	02 Dec 2021 13:31:16	U	SMPL
CCB47 - 1	CCB47	-21	0.0069	-	7470A	PPB	02 Dec 2021 13:33:33	U	SMPL