



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/3/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:35
In Date: 12/02/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 12/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117612

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4887-01	EX898	1	1.12	8.39	9.51	8.3	85.6	
M4887-02	EX899	2	1.15	8.82	9.97	8.02	77.9	
M4887-03	EX8A0	3	1.16	8.55	9.71	7.88	78.6	
M4887-04	EX8A1	4	1.15	8.80	9.95	8.15	79.5	
M4887-05	EX8A2	5	1.13	8.80	9.93	7.84	76.3	
M4887-06	EX8A3	6	1.17	8.54	9.71	7.88	78.6	
M4887-07	EX8A4	7	1.16	8.68	9.84	8.09	79.8	
M4887-10	EX8A5	8	1.19	8.57	9.76	8.00	79.5	
M4887-11	EX8A6	9	1.19	8.42	9.61	7.79	78.4	
M4887-12	EX8A7	10	1.15	8.60	9.75	8.12	81.0	
M4887-13	VHBLK001	11	1.00	1.00	2.00	2.00	100.0	VHBLK
M4887-14	EX8C8	12	1.19	8.43	9.62	7.81	78.5	
M4887-15	EX8C9	13	1.16	8.55	9.71	7.85	78.2	
M4887-16	EX8D0	14	1.17	8.77	9.94	8.61	84.8	
M4887-17	EX8D0MS	15	1.17	8.77	9.94	8.61	84.8	
M4887-18	EX8D0MSD	16	1.17	8.77	9.94	8.61	84.8	
M4887-19	EX8D1	17	1.15	8.54	9.69	7.82	78.1	
M4887-20	EX8D2	18	1.13	8.43	9.56	7.83	79.5	
M4887-21	EX8D3	19	1.19	8.32	9.51	7.62	77.3	
M4887-22	EX8D4	20	1.15	8.81	9.96	8.05	78.3	
M4887-23	EX8D5	21	1.12	8.69	9.81	7.79	76.8	
M4887-24	EX8D6	22	1.11	8.52	9.63	8.27	84.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$