



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/3/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:00
In Date: 12/02/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 12/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117615

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4888-01	EX8A8	1	1.16	8.81	9.97	8.08	78.5	
M4888-02	EX8A9	2	1.18	8.64	9.82	8.05	79.5	
M4888-03	EX8B0	3	1.18	8.32	9.5	8.83	91.9	
M4888-04	EX8B1	4	1.13	8.72	9.85	9.03	90.6	
M4888-05	EX8B2	5	1.19	8.37	9.56	8.2	83.8	
M4888-06	EX8B3	6	1.14	8.84	9.98	8.59	84.3	
M4888-07	EX8B3MS	7	1.14	8.84	9.98	8.59	84.3	
M4888-08	EX8B3MSD	8	1.14	8.84	9.98	8.59	84.3	
M4888-09	EX8B4	9	1.15	8.82	9.97	8.58	84.2	
M4888-10	EX8B5	10	1.14	8.49	9.63	8.25	83.7	
M4888-11	EX8B6	11	1.17	8.66	9.83	8.36	83.0	
M4888-12	EX8B7	12	1.1	8.67	9.77	8.01	79.7	
M4888-13	EX8B8	13	1.14	8.76	9.9	8.36	82.4	
M4888-14	EX8B9	14	1.19	8.55	9.74	8.00	79.6	
M4888-15	EX8C0	15	1.19	8.45	9.64	7.81	78.3	
M4888-16	EX8C1	16	1.19	8.71	9.9	8.17	80.1	
M4888-17	EX8C2	17	1.16	8.57	9.73	8.13	81.3	
M4888-18	EX8C3	18	1.14	8.82	9.96	8.44	82.8	
M4888-19	EX8C4	19	1.18	8.80	9.98	8.59	84.2	
M4888-20	EX8C5	20	1.15	8.60	9.75	8.67	87.4	
M4888-21	EX8C6	21	1.19	8.45	9.64	8.03	80.9	
M4888-22	EX8C7	22	1.16	8.81	9.97	8.13	79.1	
M4888-23	VHBLK001	23	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$