



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117620

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/03/2021	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/03/2021	08:16
3	CAL3	1	Water	NA	NA	20.3	10.03	12/03/2021	08:18
4	ICV	1	Water	NA	NA	20.4	6.99	12/03/2021	08:20
5	CCV1	1	Water	NA	NA	20.4	2.01	12/03/2021	08:22
6	M4920-04	1	Water	NA	NA	20.8	7.88	12/03/2021	09:50
7	M4920-04DUP	1	Water	NA	NA	20.9	7.89	12/03/2021	09:51
8	CCV2	1	Water	NA	NA	20.9	12.02	12/03/2021	10:00

WORKLIST(Hardcopy Internal Chain)

VB 117620

WorkList Name : CORROSIONITY M4920 w WorkList ID : 154828 Department : Wet-Chemistry Date : 12-03-2021 07:10:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/02/2021	Water	M4920-04	D Corrosivity	Cool 4 deg C	PARS02	D11	FDR-20211202-WC-W	12/02/2021	9040C

Date/Time 12/03/21 7:15
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: CP SM

Date/Time 12/03/21 11:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: CP SM