

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

12/3/2021 11:49

Sample id	Test id	Result	Status	Date	Errors
ICV1	Total CN	92.590	man acc	12/3/2021	
ICB1	Total CN	0.273	man acc	12/3/2021	
CCV1	Total CN	239.993	man acc	12/3/2021	
CCB1	Total CN	0.384	man acc	12/3/2021	
PB141122BL	Total CN	0.391	man acc	12/3/2021	
M4863-08	Total CN	0.417	man acc	12/3/2021	
M4863-08DUP	Total CN	0.460	man acc	12/3/2021	
CCV2	Total CN	253.904	man acc	12/3/2021	
CCB2	Total CN	0.802	man acc	12/3/2021	

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 03 10:22:36 2021

Fri Dec 03 11:39:49 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.0666	µg/l	12/3/2021 11:00:31	
5.0PPBCN	A	Total CN	P	4.9385	µg/l	12/3/2021 11:00:32	
10PPBCN	A	Total CN	P	9.3935	µg/l	12/3/2021 11:00:33	
50PPBCN	A	Total CN	P	48.781	µg/l	12/3/2021 11:00:34	
100PPBCN	A	Total CN	P	99.7052	µg/l	12/3/2021 11:00:35	
250PPBCN	A	Total CN	P	253.8433	µg/l	12/3/2021 11:00:36	
500PPBCN	A	Total CN	P	498.272	µg/l	12/3/2021 11:00:37	
ICV1	S	Total CN	P	92.5898	µg/l	12/3/2021 11:34:30	
ICB1	S	Total CN	P	0.2734	µg/l	12/3/2021 11:34:31	
CCV1	S	Total CN	P	239.9926	µg/l	12/3/2021 11:34:34	
CCB1	S	Total CN	P	0.3838	µg/l	12/3/2021 11:34:36	
PB141122BL	S	Total CN	P	0.3911	µg/l	12/3/2021 11:34:37	
M4863-08	S	Total CN	P	0.4168	µg/l	12/3/2021 11:34:38	
M4863-08DUP	S	Total CN	P	0.46	µg/l	12/3/2021 11:39:44	
CCV2	S	Total CN	P	253.9039	µg/l	12/3/2021 11:39:46	
CCB2	S	Total CN	P	0.8016	µg/l	12/3/2021 11:39:49	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

12/3/2021 11:02


Test Total CN

Accepted 12/3/2021 11:02

Factor 1219
Bias 0

Coeff. of det. 0.999904

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.0666	0.0000	-
2	5.0PPBCN	0.005	4.9385	5.0000	-1.2
3	10PPBCN	0.008	9.3935	10.0000	-6.1
4	50PPBCN	0.040	48.7810	50.0000	-2.4
5	100PPBCN	0.082	99.7052	100.0000	-0.3
6	250PPBCN	0.209	253.8433	250.0000	1.5
7	500PPBCN	0.409	498.2720	500.0000	-0.3

12 12/3/21