

LB117631


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40401



A= 0.0000e+000

B= 2.4775e-004 slope

C= -3.6610e-002 y-intercept

Rho= 0.9998631

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.015	-0.015	86	0.000	86					
0.05	0.050					205	0				-14
0.20	0.200	0.172	-0.028	841	0.0 %	841					0
2.50	2.500	2.488	-0.012	10191	0.0 %	10191					3
5.00	5.000	5.134	0.134	20869	0.0 %	20869					-1
7.50	7.500	7.449	-0.051	30215	0.0 %	30215					0
10.0	10.000	9.973	-0.027	40401	0.0 %	40401					

LB117631
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	86	-		0 SFAM01.1	PPB	12/3/2021 11:57	Std
	0.2 S01	841	-		0.2 SFAM01.1	PPB	12/3/2021 12:00	Std
	2.5 S02	10191	-		2.5 SFAM01.1	PPB	12/3/2021 12:02	Std
	5 S03	20869	-		5 SFAM01.1	PPB	12/3/2021 12:04	Std
	7.5 S04	30215	-		7.5 SFAM01.1	PPB	12/3/2021 12:06	Std
	10 S05	40401	-		10 SFAM01.1	PPB	12/3/2021 12:09	Std
ICV095	ICV095	16149	3.9643 -		SFAM01.1	PPB	12/3/2021 12:11	SMPL
ICB095	ICB095	20	-0.0317 -		SFAM01.1	PPB	12/3/2021 12:14	SMPL
CCV052	CCV052	20736	5.1007 -		SFAM01.1	PPB	12/3/2021 12:16	SMPL
CCB052	CCB052	29	-0.0294 -		SFAM01.1	PPB	12/3/2021 12:18	SMPL
PB141138BL	PBW138	37	-0.0274 -		SFAM01.1	PPB	12/3/2021 12:20	SMPL
PB141129TB	LEB129	78	-0.0173 -		SFAM01.1	PPB	12/3/2021 12:23	SMPL
M4880-01	MC0G20	52	-0.0237 -		SFAM01.1	PPB	12/3/2021 12:25	SMPL
M4880-02	MC0G21	36	-0.0277 -		SFAM01.1	PPB	12/3/2021 12:27	SMPL
M4880-03	MC0G21D	43	-0.026 -		SFAM01.1	PPB	12/3/2021 12:29	SMPL
M4880-04	MC0G21S	4364	1.0446 -		SFAM01.1	PPB	12/3/2021 12:32	SMPL
CCV053	CCV053	21074	5.1844 -		SFAM01.1	PPB	12/3/2021 12:34	SMPL
CCB053	CCB053	33	-0.0284 -		SFAM01.1	PPB	12/3/2021 12:36	SMPL