

LB117634


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52174



A= 0.0000e+000

B= 1.9242e-004

C= -3.6607e-002

Rho= 0.9999407

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/1
0.0	0.000	-0.019	-0.019	91	0.000	91					-16
0.2	0.200	0.168	-0.032	1065	0.0 %	1065					2
2.5	2.500	2.554	0.054	13462	0.0 %	13462					1
5.0	5.000	5.049	0.049	26429	0.0 %	26429					-1
7.5	7.500	7.445	-0.055	38881	0.0 %	38881					0
10.0	10.000	10.003	0.003	52174	0.0 %	52174					

LB117634 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Dec 2021 13:38:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	91	-	0.0000	7471B	PPB	03 Dec 2021 13:40:26	S	Std
0.2 - 1	50.2	1065	-	0.2000	7471B	PPB	03 Dec 2021 13:42:43	S	Std
2.5 - 1	52.5	13462	-	2.5000	7471B	PPB	03 Dec 2021 13:45:00	S	Std
5.0 - 1	55	26429	-	5.0000	7471B	PPB	03 Dec 2021 13:47:17	S	Std
7.5 - 1	57.5	38881	-	7.5000	7471B	PPB	03 Dec 2021 13:49:34	S	Std
10.0 - 1	51.1	52174	-	10.0000	7471B	PPB	03 Dec 2021 13:54:35	S	Std
ICV96 - 1	ICV96	21022	4.0085	-	7471B	PPB	03 Dec 2021 13:57:13	U	SMPL
ICB96 - 1	ICB96	-11	-0.0387	-	7471B	PPB	03 Dec 2021 13:59:27	U	SMPL
CCV54 - 1	CCV54	26885	5.1367	-	7471B	PPB	03 Dec 2021 14:01:43	U	SMPL
CCB54 - 1	CCB54	-18	-0.0401	-	7471B	PPB	03 Dec 2021 14:03:57	U	SMPL
CRA - 1	CRA	1112	0.1774	-	7471B	PPB	03 Dec 2021 14:06:13	U	SMPL
HighStd - 1	HighStd	53236	10.2073	-	7471B	PPB	03 Dec 2021 14:08:28	U	SMPL
ChkStd - 1	ChkStd	37936	7.2632	-	7471B	PPB	03 Dec 2021 14:10:43	U	SMPL
PB141155BL - 1	PBS	-54	-0.0470	-	7471B	PPB	03 Dec 2021 14:13:00	U	SMPL
PB141155BS - 1	LCSS	19865	3.7859	-	7471B	PPB	03 Dec 2021 14:15:18	U	SMPL
M4876-01 - 1	CL-02-120121	5552	1.0317	-	7471B	PPB	03 Dec 2021 14:17:35	U	SMPL
M4876-01DUP - 1	CL-02-120121DUP	5058	0.9367	-	7471B	PPB	03 Dec 2021 14:19:52	U	SMPL
M4876-01MS - 1	CL-02-120121MS	24173	4.6149	-	7471B	PPB	03 Dec 2021 14:22:09	U	SMPL
M4876-01MSD - 1	CL-02-120121MSD	24396	4.6578	-	7471B	PPB	03 Dec 2021 14:24:25	U	SMPL
M4877-01 - 1	PL-2-120121	3059	0.5520	-	7471B	PPB	03 Dec 2021 14:26:41	U	SMPL
CCV55 - 1	CCV55	27401	5.2360	-	7471B	PPB	03 Dec 2021 14:28:57	U	SMPL
CCB55 - 1	CCB55	-3	-0.0372	-	7471B	PPB	03 Dec 2021 14:31:12	U	SMPL
M4877-03 - 1	PL-5-120121	1491	0.2503	-	7471B	PPB	03 Dec 2021 14:33:29	U	SMPL
M4891-01 - 1	MINEBROOK-1-COMP	286	0.0184	-	7471B	PPB	03 Dec 2021 14:35:44	U	SMPL
M4891-03 - 1	MINEBROOK-2-COMP	1107	0.1764	-	7471B	PPB	03 Dec 2021 14:38:00	U	SMPL
M4893-01 - 1	EO-1-120121	1820	0.3136	-	7471B	PPB	03 Dec 2021 14:40:15	U	SMPL
M4910-17 - 1	SP-1-4	1025	0.1606	-	7471B	PPB	03 Dec 2021 14:42:31	U	SMPL
M4910-19 - 1	SP-1-4	1275	0.2087	-	7471B	PPB	03 Dec 2021 14:44:48	U	SMPL
M4876-01LX5 - 1		996	0.1550	-	7471B	PPB	03 Dec 2021 14:47:04	U	SMPL
M4876-01A - 1		25498	4.8698	-	7471B	PPB	03 Dec 2021 14:49:21	U	SMPL
CCV56 - 1	CCV56	27512	5.2574	-	7471B	PPB	03 Dec 2021 14:51:38	U	SMPL
CCB56 - 1	CCB56	-2	-0.0370	-	7471B	PPB	03 Dec 2021 14:53:53	U	SMPL