

LB 117635

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : 12 Instrument ID : Konelab

12/3/2021 13:29

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.241	0.0	0.078	
ICB1	0.412	0.0	0.001	
CCV1	240.956	0.0	0.198	
CCB1	0.377	0.0	0.001	
PB141125BL	0.438	0.0	0.001	
M4721-01	0.485	0.0	0.001	
M4721-01DUP	0.381	0.0	0.001	
M4721-02	0.158	0.0	0.001	
M4721-03	0.264	0.0	0.001	
CCV2	258.678	0.0	0.213	
CCB2	0.462	0.0	0.001	

N	11
Mean	54.259
SD	100.7333
CV%	185.65

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 03 13:15:38 2021

Fri Dec 03 13:20:00 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.0666	µg/l	12/3/2021 11:00:31	
5.OPPBCN	A	Total CN	P	4.9385	µg/l	12/3/2021 11:00:32	
10PPBCN	A	Total CN	P	9.3935	µg/l	12/3/2021 11:00:33	
50PPBCN	A	Total CN	P	48.781	µg/l	12/3/2021 11:00:34	
100PPBCN	A	Total CN	P	99.7052	µg/l	12/3/2021 11:00:35	
250PPBCN	A	Total CN	P	253.8433	µg/l	12/3/2021 11:00:36	
500PPBCN	A	Total CN	P	498.272	µg/l	12/3/2021 11:00:37	
ICV1	S	Total CN	P	94.2406	µg/l	12/3/2021 13:15:39	
ICB1	S	Total CN	P	0.4117	µg/l	12/3/2021 13:15:41	
CCV1	S	Total CN	P	240.9562	µg/l	12/3/2021 13:15:42	
CCB1	S	Total CN	P	0.3773	µg/l	12/3/2021 13:15:45	
PB141125ES	S	Total CN	P	0.4381	µg/l	12/3/2021 13:15:46	
M4721-01	S	Total CN	P	0.4849	µg/l	12/3/2021 13:15:47	
M4721-01IS	S	Total CN	P	0.381	µg/l	12/3/2021 13:15:48	
M4721-02	S	Total CN	P	0.1583	µg/l	12/3/2021 13:19:57	
M4721-03	S	Total CN	P	0.2635	µg/l	12/3/2021 13:19:58	
CCV2	S	Total CN	P	258.6779	µg/l	12/3/2021 13:19:59	
CCB2	S	Total CN	P	0.462	µg/l	12/3/2021 13:20:00	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

12/3/2021 11:02


Test Total CN

Accepted 12/3/2021 11:02

Factor 1219
Bias 0

Coeff. of det. 0.999904

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.0666	0.0000	-
2	5.0PPBCN	0.005	4.9385	5.0000	-1.2
3	10PPBCN	0.008	9.3935	10.0000	-6.1
4	50PPBCN	0.040	48.7810	50.0000	-2.4
5	100PPBCN	0.082	99.7052	100.0000	-0.3
6	250PPBCN	0.209	253.8433	250.0000	1.5
7	500PPBCN	0.409	498.2720	500.0000	-0.3

12 12/3/21