

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

12/3/2021 16:53

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.990	0.0	0.107	
ICB1	0.001	0.0	0.005	
CCV1	0.980	0.0	0.106	
CCB1	-0.001	0.0	0.004	
PB141101BL	0.001	0.0	0.005	
PB141101BS	0.988	0.0	0.107	
M4721-01	-0.050	0.0	-0.001	
M4721-01DUP	-0.417	0.0	-0.039	
M4721-01MS	1.006	0.0	0.109	
M4721-01MSD	1.009	0.0	0.109	
M4721-02	0.025	0.0	0.007	
M4721-03	0.014	0.0	0.006	
DOC-1	0.968	0.0	0.105	
DOC-2	0.978	0.0	0.106	
DOC-3	0.983	0.0	0.106	
DOC-4	0.987	0.0	0.107	
CCV2	0.969	0.0	0.105	
CCB2	-0.001	0.0	0.004	

N	18
Mean	0.524
SD	0.5400
CV%	103.07

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 03 16:13:11 2021

Fri Dec 03 16:52:54 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time
0.00PPM	A	Phenolics-	P	0.0029	mg/l	12/3/2021 16:13:11
0.05PPM	A	Phenolics-	P	0.0491	mg/l	12/3/2021 16:13:12
0.1PPM	A	Phenolics-	P	0.1011	mg/l	12/3/2021 16:13:13
0.25PPM	A	Phenolics-	P	0.248	mg/l	12/3/2021 16:13:14
0.50PPM	A	Phenolics-	P	0.4995	mg/l	12/3/2021 16:13:15
1.0PPM	A	Phenolics-	P	0.9983	mg/l	12/3/2021 16:13:16
2.0PPM	A	Phenolics-	P	2.0012	mg/l	12/3/2021 16:13:17
ICV1	S	Phenolics-	P	0.9898	mg/l	12/3/2021 16:45:03
ICB1	S	Phenolics-	P	0.0013	mg/l	12/3/2021 16:45:05
CCV1	S	Phenolics-	P	0.9799	mg/l	12/3/2021 16:45:07
CCB1	S	Phenolics-	P	-0.0009	mg/l	12/3/2021 16:45:10
PB141101BL	S	Phenolics-	P	0.0005	mg/l	12/3/2021 16:45:11
PB141101BS	S	Phenolics-	P	0.9883	mg/l	12/3/2021 16:45:12
M4721-01	S	Phenolics-	P	-0.0503	mg/l	12/3/2021 16:45:13
M4721-01DUP	S	Phenolics-	P	-0.4167	mg/l	12/3/2021 16:45:14
M4721-01MS	S	Phenolics-	P	1.0062	mg/l	12/3/2021 16:52:45
M4721-01MSD	S	Phenolics-	P	1.0094	mg/l	12/3/2021 16:52:46
M4721-02	S	Phenolics-	P	0.0251	mg/l	12/3/2021 16:52:47
M4721-03	S	Phenolics-	P	0.0139	mg/l	12/3/2021 16:52:48
DOC-1	S	Phenolics-	P	0.9682	mg/l	12/3/2021 16:52:49
DOC-2	S	Phenolics-	P	0.9779	mg/l	12/3/2021 16:52:50
DOC-3	S	Phenolics-	P	0.9829	mg/l	12/3/2021 16:52:51
DOC-4	S	Phenolics-	P	0.9872	mg/l	12/3/2021 16:52:52
CCV2	S	Phenolics-	P	0.9692	mg/l	12/3/2021 16:52:53
CCB2	S	Phenolics-	P	-0.0009	mg/l	12/3/2021 16:52:54

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

12/3/2021 16:14


Test Phenolics-

Accepted 12/3/2021 16:14

Factor 9.668
Bias 0.004

Coeff. of det. 0.999994

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.005	0.0029	0.0000	-
2	Phenol-2	0.010	0.0491	0.0500	-1.8
3	Phenol-2	0.015	0.1011	0.1000	1.1
4	Phenol-2	0.030	0.2480	0.2500	-0.9
5	Phenol-2	0.056	0.4995	0.5000	-0.1
6	Phenol-2	0.108	0.9983	1.0000	-0.2
7	Phenol-2	0.211	2.0012	2.0000	0.1

12 12/03/21