


LB117648

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

36791



Linear

A= 0.0000e+000

B= 2.6785e-004 slope

C= 5.9746e-003 y-intercept

Rho= 0.9995963

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.020	0.020	52	0.000	52					=
0.05	0.050					205	0				-2
0.20	0.200	0.196	-0.004	711	0.0 %	711					-4
2.50	2.500	2.406	-0.094	8961	0.0 %	8961					1
5.00	5.000	5.025	0.025	18739	0.0 %	18739					3
7.50	7.500	7.692	0.192	28693	0.0 %	28693					-1
10.0	10.000	9.861	-0.139	36791	0.0 %	36791					

LB117648
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	52	-		0 SFAM01.1	PPB	12/4/2021 9:27	Std
	0.2 S01	711	-		0.2 SFAM01.1	PPB	12/4/2021 9:29	Std
	2.5 S02	8961	-		2.5 SFAM01.1	PPB	12/4/2021 9:31	Std
	5 S03	18739	-		5 SFAM01.1	PPB	12/4/2021 9:34	Std
	7.5 S04	28693	-		7.5 SFAM01.1	PPB	12/4/2021 9:36	Std
	10 S05	36791	-		10 SFAM01.1	PPB	12/4/2021 9:41	Std
ICV096	ICV096	14960	4.0131 -		SFAM01.1	PPB	12/4/2021 9:43	SMPL
ICB096	ICB096	-19	0.0009 -		SFAM01.1	PPB	12/4/2021 9:46	SMPL
CCV057	CCV057	18742	5.0261 -		SFAM01.1	PPB	12/4/2021 9:51	SMPL
CCB057	CCB057	0	0.006 -		SFAM01.1	PPB	12/4/2021 9:53	SMPL
PB141171BL	PBW171	29	0.0137 -		SFAM01.1	PPB	12/4/2021 9:58	SMPL
M4941-01	JDGG4	332	0.0949 -		SFAM01.1	PPB	12/4/2021 10:00	SMPL
M4941-02	JDGG5	294	0.0847 -		SFAM01.1	PPB	12/4/2021 10:02	SMPL
M4941-03	JDGG5D	288	0.0831 -		SFAM01.1	PPB	12/4/2021 10:04	SMPL
M4941-04	JDGG5S	4549	1.2244 -		SFAM01.1	PPB	12/4/2021 10:07	SMPL
CCV058	CCV058	20266	5.4343 -		SFAM01.1	PPB	12/4/2021 10:09	SMPL
CCB058	CCB058	17	0.0105 -		SFAM01.1	PPB	12/4/2021 10:11	SMPL