


LB117657

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.9824e-004

C= -7.5341e-002

Rho= 0.9998768

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.020	-0.020	281	0.000	281					
0.2	0.200	0.169	-0.031	1232	0.0 %	1232					
2.5	2.500	2.522	0.022	13102	0.0 %	13102					
5.0	5.000	5.110	0.110	26158	0.0 %	26158					
7.5	7.500	7.422	-0.078	37818	0.0 %	37818					
10.0	10.000	9.997	-0.003	50806	0.0 %	50806					

%D
= 16
1
2
-1
0

LB117657 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Dec 2021 09:14:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	281	-	0.0000	7471B	PPB	06 Dec 2021 10:01:44	S	Std
0.2 - 1	58.2	1232	-	0.2000	7471B	PPB	06 Dec 2021 10:04:01	S	Std
2.5 - 1	52.5	13102	-	2.5000	7471B	PPB	06 Dec 2021 10:06:18	S	Std
5.0 - 1	55	26158	-	5.0000	7471B	PPB	06 Dec 2021 10:08:34	S	Std
7.5 - 1	59.5	37818	-	7.5000	7471B	PPB	06 Dec 2021 10:10:51	S	Std
10.0 - 1	51	50806	-	10.0000	7471B	PPB	06 Dec 2021 10:18:28	S	Std
ICV97 - 1	ICV97	20442	3.9771	-	7471B	PPB	06 Dec 2021 10:21:20	U	SMPL
ICB97 - 1	ICB97	375	-0.0010	-	7471B	PPB	06 Dec 2021 10:23:35	U	SMPL
CCV59 - 1	CCV59	26490	5.1761	-	7471B	PPB	06 Dec 2021 10:25:50	U	SMPL
CCB59 - 1	CCB59	389	0.0018	-	7471B	PPB	06 Dec 2021 10:28:05	U	SMPL
CRA - 1	CRA	1349	0.1921	-	7471B	PPB	06 Dec 2021 10:41:17	U	SMPL
HighStd - 1	HighStd	51444	10.1231	-	7471B	PPB	06 Dec 2021 10:43:32	U	SMPL
ChkStd - 1	ChkStd	36731	7.2063	-	7471B	PPB	06 Dec 2021 10:45:47	U	SMPL
PB141168BL - 1	PBS	132	-0.0492	-	7471B	PPB	06 Dec 2021 10:48:05	U	SMPL
PB141168BS - 1	LCSS	20896	4.0671	-	7471B	PPB	06 Dec 2021 10:50:21	U	SMPL
M4909-01 - 1	HG-SOIL-PILE	9936	1.8944	-	7471B	PPB	06 Dec 2021 10:57:27	U	SMPL
M4909-01DUP - 1	HG-SOIL-PILEDUP	8361	1.5822	-	7471B	PPB	06 Dec 2021 10:59:43	U	SMPL
M4909-01MS - 1	HG-SOIL-PILEMS	27384	5.3533	-	7471B	PPB	06 Dec 2021 11:05:06	U	SMPL
M4909-01MSD - 1	HG-SOIL-PILEMSD	27061	5.2893	-	7471B	PPB	06 Dec 2021 11:07:22	U	SMPL
M4915-01 - 1	B-1-7	29798	5.8319	-	7471B	PPB	06 Dec 2021 11:09:38	U	SMPL
CCV60 - 1	CCV60	26467	5.1716	-	7471B	PPB	06 Dec 2021 11:11:53	U	SMPL
CCB60 - 1	CCB60	397	0.0034	-	7471B	PPB	06 Dec 2021 11:14:09	U	SMPL
M4915-03 - 1	B-2-10-15	200290	39.6307	-	7471B	PPB	06 Dec 2021 11:16:25	U	SMPL
M4915-06 - 1	B-3-13	57082	11.2408	-	7471B	PPB	06 Dec 2021 11:18:41	U	SMPL
M4915-08 - 1	B-4-7	6456	1.2045	-	7471B	PPB	06 Dec 2021 11:20:59	U	SMPL
M4915-10 - 1	B-5-4	21072	4.1020	-	7471B	PPB	06 Dec 2021 11:23:17	U	SMPL
M4915-13 - 1	B-6-4	57917	11.4063	-	7471B	PPB	06 Dec 2021 11:25:33	U	SMPL
M4920-01 - 1	FDR-20211202-WC-S	41087	8.0699	-	7471B	PPB	06 Dec 2021 11:27:49	U	SMPL
M4932-01 - 1	DRUM1-2	4996	0.9151	-	7471B	PPB	06 Dec 2021 11:32:07	U	SMPL
M4939-01 - 1	OR-03-120321	2470	0.4143	-	7471B	PPB	06 Dec 2021 11:34:24	U	SMPL
PB141169BL - 1	PBS	243	-0.0272	-	7471B	PPB	06 Dec 2021 11:36:40	U	SMPL
PB141169BS - 1	LCSS	19953	3.8802	-	7471B	PPB	06 Dec 2021 11:38:56	U	SMPL
CCV61 - 1	CCV61	26548	5.1876	-	7471B	PPB	06 Dec 2021 11:41:11	U	SMPL
CCB61 - 1	CCB61	406	0.0051	-	7471B	PPB	06 Dec 2021 11:43:27	U	SMPL
M4927-01 - 1	PH1-BOT-3	982	0.1193	-	7471B	PPB	06 Dec 2021 11:45:42	U	SMPL
M4927-01DUP - 1	PH1-BOT-3DUP	1010	0.1249	-	7471B	PPB	06 Dec 2021 11:50:46	U	SMPL
M4927-01MS - 1	PH1-BOT-3MS	20810	4.0501	-	7471B	PPB	06 Dec 2021 11:53:01	U	SMPL
M4927-01MSD - 1	PH1-BOT-3MSD	21252	4.1377	-	7471B	PPB	06 Dec 2021 11:55:17	U	SMPL
M4927-02 - 1	PH1-BOT-4	1305	0.1834	-	7471B	PPB	06 Dec 2021 11:57:32	U	SMPL
M4927-03 - 1	PH1-BOT-5	853	0.0938	-	7471B	PPB	06 Dec 2021 11:59:49	U	SMPL
M4927-04 - 1	PH1-BOT-6	1302	0.1828	-	7471B	PPB	06 Dec 2021 12:02:05	U	SMPL
M4927-05 - 1	PH1-BOT-7	1649	0.2516	-	7471B	PPB	06 Dec 2021 12:04:22	U	SMPL
M4927-06 - 1	PH1-BOT-8	14091	2.7181	-	7471B	PPB	06 Dec 2021 12:06:40	U	SMPL
M4927-07 - 1	PH1-BOT-9	1447	0.2115	-	7471B	PPB	06 Dec 2021 12:08:56	U	SMPL
CCV62 - 1	CCV62	26494	5.1769	-	7471B	PPB	06 Dec 2021 12:11:12	U	SMPL
CCB62 - 1	CCB62	379	-0.0002	-	7471B	PPB	06 Dec 2021 12:13:27	U	SMPL
M4927-08 - 1	PH1-BOT-15	1245	0.1715	-	7471B	PPB	06 Dec 2021 12:15:43	U	SMPL
M4927-09 - 1	PH1-BOT-10	3554	0.6292	-	7471B	PPB	06 Dec 2021 12:17:58	U	SMPL
M4927-10 - 1	PH1-BOT-11	1015	0.1259	-	7471B	PPB	06 Dec 2021 12:20:14	U	SMPL
M4927-11 - 1	PH1-BOT-12	910	0.1051	-	7471B	PPB	06 Dec 2021 12:22:30	U	SMPL
M4927-12 - 1	PH1-BOT-13	1364	0.1951	-	7471B	PPB	06 Dec 2021 12:24:45	U	SMPL
M4927-13 - 1	PH1-BOT-14	1023	0.1275	-	7471B	PPB	06 Dec 2021 12:27:02	U	SMPL
M4927-14 - 1	PH1-BOT-20	11965	2.2966	-	7471B	PPB	06 Dec 2021 12:29:18	U	SMPL
CCV63 - 1	CCV63	26955	5.2683	-	7471B	PPB	06 Dec 2021 12:31:35	U	SMPL
CCB63 - 1	CCB63	381	0.0002	-	7471B	PPB	06 Dec 2021 12:33:52	U	SMPL
M4909-01LX5 - 1		2044	0.3299	-	7471B	PPB	06 Dec 2021 12:36:08	U	SMPL
M4909-01A - 1		28985	5.6707	-	7471B	PPB	06 Dec 2021 12:38:23	U	SMPL
M4927-01LX5 - 1		400	0.0040	-	7471B	PPB	06 Dec 2021 12:40:39	U	SMPL
M4927-01A - 1		21281	4.1435	-	7471B	PPB	06 Dec 2021 12:42:54	U	SMPL
M4915-03DLX10 - 1		21620	4.2107	-	7471B	PPB	06 Dec 2021 12:45:10	U	SMPL
M4915-06DLX2 - 1		28625	5.5994	-	7471B	PPB	06 Dec 2021 12:47:26	U	SMPL
M4915-13DLX2 - 1		29441	5.7611	-	7471B	PPB	06 Dec 2021 12:49:41	U	SMPL
CCV64 - 1	CCV64	26803	5.2382	-	7471B	PPB	06 Dec 2021 12:57:07	U	SMPL
CCB64 - 1	CCB64	366	-0.0028	-	7471B	PPB	06 Dec 2021 12:59:23	U	SMPL