



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117661

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/06/2021	12:10
2	CAL2	1	Water	NA	NA	20.3	7.00	12/06/2021	12:11
3	CAL3	1	Water	NA	NA	20.4	10.03	12/06/2021	12:12
4	ICV	1	Water	NA	NA	20.3	7.02	12/06/2021	12:15
5	CCV1	1	Water	NA	NA	20.3	2.01	12/06/2021	12:16
6	M4934-05	1	Water	NA	NA	20.8	11.59	12/06/2021	12:45
7	M4934-06	1	Water	NA	NA	20.9	6.57	12/06/2021	12:47
8	M4934-06DUP	1	Water	NA	NA	20.8	6.58	12/06/2021	12:48
9	M4934-07	1	Water	NA	NA	21.3	9.17	12/06/2021	12:50
10	CCV2	1	Water	NA	NA	20.5	12.03	12/06/2021	13:00

117661

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M4934 w WorkList ID : 154903 Department : Wet-Chemistry Date : 12-06-2021 11:51:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/03/2021	Water	M4934-05	Corrosivity	Cool 4 deg C	PSEG03	F42	17078	12/03/2021	9040C
12/03/2021	Water	M4934-06	Corrosivity	Cool 4 deg C	PSEG03	F42	17094	12/03/2021	9040C
12/03/2021	Water	M4934-07	Corrosivity	Cool 4 deg C	PSEG03	F42	17097	12/03/2021	9040C

Date/Time 12/06/21 12:00
Raw Sample Received by: JP (w/c)
Raw Sample Relinquished by: JP (w/c)

Date/Time 12/06/21 12:00
Raw Sample Received by: JP (w/c)
Raw Sample Relinquished by: JP (w/c)